



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111553

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/16/2020	12:25
2	CAL2	1	Water	NA	NA	20.1	7.00	10/16/2020	12:26
3	CAL3	1	Water	NA	NA	20.1	10.03	10/16/2020	12:27
4	ICV	1	Water	NA	NA	20.3	7.00	10/16/2020	12:30
5	CCV1	1	Water	NA	NA	20.1	2.02	10/16/2020	12:31
6	L4438-01	1	Water	NA	NA	14.7	4.69	10/16/2020	12:33
7	L4438-01DUP	1	Water	NA	NA	14.8	4.70	10/16/2020	12:34
8	L4438-02	1	Water	NA	NA	16.0	4.61	10/16/2020	12:35
9	L4438-03	1	Water	NA	NA	16.5	4.86	10/16/2020	12:36
10	CCV2	1	Water	NA	NA	20.1	12.02	10/16/2020	12:37

LB 111553

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4438 WorkList ID : 143402 Department : Wet-Chemistry Date : 10-16-2020 12:15:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/15/2020	Water	L4438-01	pH	Cool 4 deg C	STER09	P12	CELL-A -POOL-WATER-(DE-IC)	10/15/2020	9040C
10/15/2020	Water	L4438-02	pH	Cool 4 deg C	STER09	P12	CELL-B-C-POOL-WATER-(DE)	10/15/2020	9040C
10/15/2020	Water	L4438-03	pH	Cool 4 deg C	STER09	P12	CELL-D-POOL-WATER-(DE-IC)	10/15/2020	9040C

Date/Time 10/16/2020 12:20
 Received by: J.P. (Wet-Chem)
 Relinquished by: AM (Wet-Chem)

Date/Time 10-16-2020 13:20
 Received by: J.P. (Wet-Chem)
 Relinquished by: AM (Wet-Chem)