



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111554

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/16/2020	12:30
2	CAL2	1	Water	NA	NA	20.1	7.00	10/16/2020	12:31
3	CAL3	1	Water	NA	NA	20.3	10.02	10/16/2020	12:34
4	ICV	1	Water	NA	NA	20.2	7.00	10/16/2020	12:35
5	CCV1	1	Water	NA	NA	20.2	2.01	10/16/2020	12:36
6	L4411-07	1	Solid	20.03	20	22.1	6.69	10/16/2020	12:37
7	L4411-07DUP	1	Solid	20.04	20	22.2	6.70	10/16/2020	12:38
8	CCV2	1	Water	NA	NA	20.2	12.02	10/16/2020	12:40

VB 111554

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L411 S WorkList ID : 143403 Department : Wet-Chemistry Date : 10-16-2020 12:15:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/15/2020	Solid	L4411-07	pH	Cool 4 deg C	PSEG01	O12	32697	10/15/2020	9045D

Date/Time 10/16/2020 12:15:20
Received by: JB / wet
Relinquished by: JB / wet (10m)

Date/Time 10/16/2020 13:25
Received by: JB / wet
Relinquished by: JB / wet (10m)