



Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB111556
Analysis Date: 10/16/2020
BalanceID: WC SC-3
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: apatel
Extraction Date: 10/16/2020
Extraction IN Time: 08:35
Extraction OUT Time: 09:00

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB111556BL	LB111556BL	WATER	1.3	1000	100	2.4135	2.4135	3.02	2.4135	2.4135	0.0000	0
2	LB111556BS	LB111556BS	WATER	1.3	1000	100	2.9674	2.9674	3.01	2.9862	2.9862	0.0188	18.8
3	L4392-17	HACK-MH	WATER	1.6	1000	100	2.9617	2.9617	3.03	2.9623	2.9623	0.0006	0.6
4	L4405-01	GRAB	WATER	1.6	1000	100	2.9519	2.9519	3.02	2.9834	2.9834	0.0315	31.5
5	L4412-06	98-02	WATER	1.6	1000	100	3.0152	3.0152	3.03	3.0234	3.0234	0.0082	8.2
6	L4427-01	STORM-DRAIN-1	WATER	1.3	1000	100	3.0514	3.0514	3.04	3.0516	3.0516	0.0002	0.2
7	L4427-01MS	STORM-DRAIN-1	WATER	1.3	1000	100	3.1566	3.1566	3.05	3.1768	3.1768	0.0202	20.2
8	L4427-02	STORM-DRAIN-2	WATER	1.3	1000	100	3.0374	3.0374	3.04	3.0396	3.0396	0.0022	2.2
9	L4427-02DUP	STORM-DRAIN-2DUP	WATER	1.3	1000	100	3.1957	3.1957	3.05	3.1978	3.1978	0.0021	2.1



QC Batch# LB111556 Test: TPH Analysis Date: 10/16/2020

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W2749
PH STRIPS BOX-0.2.5	M4588
Sodium Sulfate	EP2088
1:1 HCL	WP85192
Silica Gel	W2546
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	10.00 ML	WP85193
LCSWD	NA	NA
MS/MSD	10.00 ML	WP85194

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-3

Before Analysis

0.0020 gram Balance:	<u>0.0019</u>	In OVEN TEMP1 :	<u>70 °C</u>	Dessicator Time In1	<u>10:16</u>
1.0000 gram Balance:	<u>1.0004</u>	In Time1:	<u>09:15</u>		
Bal Check Time:	<u>08:40</u>	Out OVEN TEMP1:	<u>70 °C</u>	Dessicator Time Out1	<u>10:45</u>
		Out Time1	<u>10:15</u>		

After Analysis

0.0020 gram Balance:	<u>0.0021</u>	In OVEN TEMP2 :	<u>71 °C</u>	Dessicator Time In2	<u>11:42</u>
1.0000 gram Balance:	<u>1.0005</u>	In Time2:	<u>11:15</u>		
Bal Check Time:	<u>12:45</u>	Out OVEN TEMP2:	<u>71 °C</u>	Dessicator Time Out2	<u>12:15</u>
		Out Time2:	<u>11:40</u>		

111556

WORKLIST(Hardcopy Internal Chain)

WorkList Name : tph 14427 WorkList ID : 143381 Department : Wet-Chemistry Date : 10-16-2020 08:32:39

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/21/2020	Water	L4392-17	TPH	Conc H2SO4	PSEG01	O12	HACK-MH	10/14/2020	1664A
10/27/2020	Water	L4405-01	TPH	Conc H2SO4	ARAM01	P11	GRAB	10/14/2020	1664A
10/22/2020	Water	L4412-06	TPH	Conc H2SO4	PSEG01	O12	98-02	10/15/2020	1664A
10/29/2020	Water	L4427-01	TPH	Conc H2SO4	KWAR01	P13	STORM-DRAIN-1	10/15/2020	1664A
10/29/2020	Water	L4427-02	TPH	Conc H2SO4	KWAR01	P13	STORM-DRAIN-2	10/15/2020	1664A

Date/Time 10-16-2020 8:33
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 10-16-2020 11:00
 Received by: [Signature]
 Relinquished by: [Signature]