


LB111560

7471B

INSTRUMENT 1321CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.046	0.046	31	0.000	31				
0.2	0.200	0.231	0.031	931	0.0 %	931				
2.5	2.500	2.423	-0.077	11614	0.0 %	11614				
5.0	5.000	5.024	0.024	24287	0.0 %	24287				
7.5	7.500	7.390	-0.110	35818	0.0 %	35818				
10.0	10.000	10.086	0.086	48955	0.0 %	48955				

LB111560 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 16 Oct 2020 14:36:23

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	31	-	0.0000	7471B	PPB	16 Oct 2020 14:36:39	S	Std
0.2 - 1	50.2	931	-	0.2000	7471B	PPB	16 Oct 2020 14:38:56	S	Std
2.5 - 1	52.5	11614	-	2.5000	7471B	PPB	16 Oct 2020 14:41:13	S	Std
5.0 - 1	55	24287	-	5.0000	7471B	PPB	16 Oct 2020 14:43:30	S	Std
7.5 - 1	57.5	35818	-	7.5000	7471B	PPB	16 Oct 2020 14:45:46	S	Std
10.0 - 1	510	48955	-	10.0000	7471B	PPB	16 Oct 2020 14:50:40	S	Std
ICV61 - 1	ICV61	18824	3.9027	-	7471B	PPB	16 Oct 2020 14:53:22	U	SMPL
ICB61 - 1	ICB61	-115	0.0161	-	7471B	PPB	16 Oct 2020 14:55:37	U	SMPL
CCV44 - 1	CCV44	24715	5.1117	-	7471B	PPB	16 Oct 2020 14:57:52	U	SMPL
CCB44 - 1	CCB44	-122	0.0146	-	7471B	PPB	16 Oct 2020 15:00:07	U	SMPL
CRA - 1	CRA	905	0.2254	-	7471B	PPB	16 Oct 2020 15:02:24	U	SMPL
HighStd - 1	HighStd	48490	9.9907	-	7471B	PPB	16 Oct 2020 15:04:39	U	SMPL
ChkStd - 1	ChkStd	34500	7.1197	-	7471B	PPB	16 Oct 2020 15:06:54	U	SMPL
PB132440BL - 1	PBS	-133	0.0124	-	7471B	PPB	16 Oct 2020 15:09:12	U	SMPL
PB132440BS - 1	LCSS	17729	3.6780	-	7471B	PPB	16 Oct 2020 15:11:29	U	SMPL
L4214-03 - 1	MDL-SOIL-03-QT4-2020	101	0.0604	-	7471B	PPB	16 Oct 2020 15:18:58	U	SMPL
L4218-01 - 1	HD-01-101520	1497	0.3469	-	7471B	PPB	16 Oct 2020 15:21:15	U	SMPL
L4219-09 - 1	CL-02-101520	1951	0.4400	-	7471B	PPB	16 Oct 2020 15:23:31	U	SMPL
L4224-06 - 1	OR-03-101520	4848	1.0346	-	7471B	PPB	16 Oct 2020 15:33:48	U	SMPL
L4224-06DUP - 1	OR-03-101520DUP	4638	0.9915	-	7471B	PPB	16 Oct 2020 15:36:02	U	SMPL
CCV45 - 1	CCV45	24374	5.0417	-	7471B	PPB	16 Oct 2020 15:43:01	U	SMPL
CCB45 - 1	CCB45	-70	0.0253	-	7471B	PPB	16 Oct 2020 15:45:15	U	SMPL
L4224-06MS - 1	OR-03-101520MS	27305	5.6432	-	7471B	PPB	16 Oct 2020 15:47:31	U	SMPL
L4224-06MSD - 1	OR-03-101520MSD	26364	5.4501	-	7471B	PPB	16 Oct 2020 15:49:46	U	SMPL
L4224-06LX5 - 1		802	0.2043	-	7471B	PPB	16 Oct 2020 15:52:02	U	SMPL
L4224-06A - 1		26484	5.4747	-	7471B	PPB	16 Oct 2020 15:54:18	U	SMPL
CCV46 - 1	CCV46	25761	5.3263	-	7471B	PPB	16 Oct 2020 15:56:34	U	SMPL
CCB46 - 1	CCB46	-87	0.0218	-	7471B	PPB	16 Oct 2020 15:58:51	U	SMPL