



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111569

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	10/16/2020	16:20
2	CAL2	1	Water	NA	NA	21.1	7.00	10/16/2020	16:21
3	CAL3	1	Water	NA	NA	21.1	10.03	10/16/2020	16:22
4	ICV	1	Water	NA	NA	21.2	7.01	10/16/2020	16:24
5	CCV1	1	Water	NA	NA	21.1	2.02	10/16/2020	16:25
6	L4434-03	1	Water	NA	NA	18.2	6.57	10/16/2020	16:26
7	L4434-04	1	Water	NA	NA	18.3	7.77	10/16/2020	16:27
8	L4434-05	1	Water	NA	NA	19.0	5.25	10/16/2020	16:30
9	L4434-06	1	Water	NA	NA	19.2	4.70	10/16/2020	16:32
10	L4434-06DUP	1	Water	NA	NA	19.2	4.71	10/16/2020	16:34
11	CCV2	1	Water	NA	NA	21.3	12.02	10/16/2020	16:35

WORKLIST(Hardcopy Internal Chain)

287 111569

WorkList Name : ph 14434 w WorkList ID : 143413 Department : Wet-Chemistry Date : 10-16-2020 16:06:15

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/16/2020	Water	L4434-03	pH	Cool 4 deg C	PSEG01	O21	40978-A	10/16/2020	9040C
10/16/2020	Water	L4434-04	pH	Cool 4 deg C	PSEG01	O21	16757-A	10/16/2020	9040C
10/16/2020	Water	L4434-05	pH	Cool 4 deg C	PSEG01	O21	16743	10/16/2020	9040C
10/16/2020	Water	L4434-06	pH	Cool 4 deg C	PSEG01	O21	16804	10/16/2020	9040C

Date/Time 10/16/2020 16:10
 Received by: JO (sm)
 Relinquished by: JO (sm)

Date/Time 10/16/2020 17:30
 Received by: CP (sm)
 Relinquished by: JO (sm)