



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111570

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	10/16/2020	16:20
2	CAL2	1	Water	NA	NA	21.1	7.00	10/16/2020	16:21
3	CAL3	1	Water	NA	NA	21.1	10.03	10/16/2020	16:22
4	ICV	1	Water	NA	NA	21.2	7.01	10/16/2020	16:24
5	CCV1	1	Water	NA	NA	21.1	2.02	10/16/2020	16:25
6	L4434-01	1	Solid	20.03	20	20.7	6.89	10/16/2020	16:26
7	L4434-02	1	Solid	20.04	20	20.6	8.05	10/16/2020	16:28
8	L4434-02DUP	1	Solid	20.03	20	20.7	8.06	10/16/2020	16:29
9	CCV2	1	Water	NA	NA	20.2	12.02	10/16/2020	16:30

WORKLIST(Hardcopy Internal Chain)

4311570

WorkList Name : ph 14434 s WorkList ID : 143414 Department : Wet-Chemistry Date : 10-16-2020 16:07:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/16/2020	Solid	L4434-01	pH	Cool 4 deg C	PSEG01	O21	40978-S	10/16/2020	9045D
10/16/2020	Solid	L4434-02	pH	Cool 4 deg C	PSEG01	O21	16757-S	10/16/2020	9045D

Date/Time 10-16-2020 16:10
 Received by: 19 (wet Chem)
 Relinquished by: CL (sm)

Date/Time 10-16-2020 17:30
 Received by: CL (sm)
 Relinquished by: 19