


LB111581
INSTRUMENT ID: CV1

7470A

Linear



A= 0.0000e+000

B= 2.6402e-004

C= -1.1586e-002

Rho= 0.9996439

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.005	-0.005	24	0.000	24				
0.2	0.200	0.194	-0.006	780	0.0 %	780				
2.5	2.500	2.395	-0.105	9115	0.0 %	9115				
5.0	5.000	5.136	0.136	19498	0.0 %	19498				
7.5	7.500	7.609	0.109	28862	0.0 %	28862				
10.0	10.000	9.871	-0.129	37430	0.0 %	37430				

LB111581 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Oct 2020 09:14:31

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	24	-	0.0000	7470A	PPB	19 Oct 2020 09:27:51	S	Std
0.2 - 1	50.2	780	-	0.2000	7470A	PPB	19 Oct 2020 09:30:08	S	Std
2.5 - 1	52.5	9115	-	2.5000	7470A	PPB	19 Oct 2020 09:32:25	S	Std
5.0 - 1	55	19498	-	5.0000	7470A	PPB	19 Oct 2020 09:34:42	S	Std
7.5 - 1	57.5	28862	-	7.5000	7470A	PPB	19 Oct 2020 09:36:58	S	Std
10.0 - 1	518	37430	-	10.0000	7470A	PPB	19 Oct 2020 09:39:14	S	Std
ICV62 - 1	ICV62	14352	3.7777	-	7470A	PPB	19 Oct 2020 09:42:01	U	SMPL
ICB62 - 1	ICB62	-84	-0.0338	-	7470A	PPB	19 Oct 2020 09:44:16	U	SMPL
CCV47 - 1	CCV47	18527	4.8800	-	7470A	PPB	19 Oct 2020 09:46:31	U	SMPL
CCB47 - 1	CCB47	-77	-0.0319	-	7470A	PPB	19 Oct 2020 09:48:46	U	SMPL
CRA - 1	CRA	755	0.1878	-	7470A	PPB	19 Oct 2020 09:51:01	U	SMPL
HighStd - 1	HighStd	37529	9.8970	-	7470A	PPB	19 Oct 2020 09:53:15	U	SMPL
ChkStd - 1	ChkStd	26311	6.9352	-	7470A	PPB	19 Oct 2020 09:55:30	U	SMPL
PB132421BL - 1	PBW	-122	-0.0438	-	7470A	PPB	19 Oct 2020 09:57:47	U	SMPL
PB132421BS - 1	LCSW	13969	3.6766	-	7470A	PPB	19 Oct 2020 10:00:03	U	SMPL
L4392-18 - 1	HACK-MH	-83	-0.0335	-	7470A	PPB	19 Oct 2020 10:02:19	U	SMPL
L4392-18DUP - 1	HACK-MHDUP	21	-0.0060	-	7470A	PPB	19 Oct 2020 10:04:36	U	SMPL
L4392-18MS - 1	HACK-MHMS	14261	3.7537	-	7470A	PPB	19 Oct 2020 10:06:53	U	SMPL
L4392-18MSD - 1	HACK-MHMSD	14615	3.8471	-	7470A	PPB	19 Oct 2020 10:09:10	U	SMPL
L4412-07 - 1	98-02	-25	-0.0182	-	7470A	PPB	19 Oct 2020 10:11:26	U	SMPL
CCV48 - 1	CCV48	18770	4.9442	-	7470A	PPB	19 Oct 2020 10:13:41	U	SMPL
CCB48 - 1	CCB48	-65	-0.0287	-	7470A	PPB	19 Oct 2020 10:15:56	U	SMPL
PB132434BL - 1	PBW	7	-0.0097	-	7470A	PPB	19 Oct 2020 10:18:11	U	SMPL
PB132434BS - 1	LCSW	13895	3.6570	-	7470A	PPB	19 Oct 2020 10:20:26	U	SMPL
L4218-04 - 1	HD-01-101520	-35	-0.0208	-	7470A	PPB	19 Oct 2020 10:22:42	U	SMPL
L4218-04DUP - 1	HD-01-101520DUP	21	-0.0060	-	7470A	PPB	19 Oct 2020 10:24:57	U	SMPL
L4218-04MS - 1	HD-01-101520MS	14765	3.8867	-	7470A	PPB	19 Oct 2020 10:27:13	U	SMPL
L4218-04MSD - 1	HD-01-101520MSD	15378	4.0486	-	7470A	PPB	19 Oct 2020 10:29:29	U	SMPL
L4219-16 - 1	CL-02-101520	-35	-0.0208	-	7470A	PPB	19 Oct 2020 10:31:46	U	SMPL
L4223-08 - 1	PL-01-101420	25	-0.0050	-	7470A	PPB	19 Oct 2020 10:34:03	U	SMPL
L4224-10 - 1	OR-03-101520	24	-0.0052	-	7470A	PPB	19 Oct 2020 10:36:19	U	SMPL
L4227-04 - 1	BU-01-101420	16	-0.0074	-	7470A	PPB	19 Oct 2020 10:38:35	U	SMPL
CCV49 - 1	CCV49	18254	4.8079	-	7470A	PPB	19 Oct 2020 10:40:51	U	SMPL
CCB49 - 1	CCB49	-51	-0.0251	-	7470A	PPB	19 Oct 2020 10:43:07	U	SMPL
L4392-21 - 1	TP-1	-4	-0.0126	-	7470A	PPB	19 Oct 2020 10:45:22	U	SMPL
L4392-22 - 1	TP-2	90	0.0122	-	7470A	PPB	19 Oct 2020 10:47:38	U	SMPL
L4392-23 - 1	TP-3	45	0.0003	-	7470A	PPB	19 Oct 2020 10:49:53	U	SMPL
L4392-24 - 1	TP-4	46	0.0006	-	7470A	PPB	19 Oct 2020 10:52:09	U	SMPL
L4404-02 - 1	SHELL-L15-L14-NORTH	41	-0.0008	-	7470A	PPB	19 Oct 2020 10:54:25	U	SMPL
L4406-08 - 1	JC-02-101420-A	55	0.0029	-	7470A	PPB	19 Oct 2020 10:56:41	U	SMPL
L4406-13 - 1	JC-02-101420-B	47	0.0008	-	7470A	PPB	19 Oct 2020 10:58:58	U	SMPL
L4413-01 - 1	30688	51	0.0019	-	7470A	PPB	19 Oct 2020 11:01:14	U	SMPL
L4418-01 - 1	PB-2020-WS-000-02	50	0.0016	-	7470A	PPB	19 Oct 2020 11:03:31	U	SMPL
L4432-01 - 1	WCS-DD-G-01	80	0.0095	-	7470A	PPB	19 Oct 2020 11:05:47	U	SMPL
CCV50 - 1	CCV50	18444	4.8581	-	7470A	PPB	19 Oct 2020 11:08:03	U	SMPL
CCB50 - 1	CCB50	-47	-0.0240	-	7470A	PPB	19 Oct 2020 11:10:18	U	SMPL
PB132436BL - 1	PBW	23	-0.0055	-	7470A	PPB	19 Oct 2020 11:12:34	U	SMPL
PB132436BS - 1	LCSW	13736	3.6151	-	7470A	PPB	19 Oct 2020 11:14:49	U	SMPL
L4412-06 - 1	98-02	323	0.0737	-	7470A	PPB	19 Oct 2020 11:17:05	U	SMPL
L4412-06DUP - 1	98-02DUP	183	0.0367	-	7470A	PPB	19 Oct 2020 11:19:21	U	SMPL
L4412-06MS - 1	98-02MS	16142	4.2503	-	7470A	PPB	19 Oct 2020 11:21:36	U	SMPL
L4412-06MSD - 1	98-02MSD	14901	3.9226	-	7470A	PPB	19 Oct 2020 11:23:53	U	SMPL
PB132421TB - 1	PB132425TB	-108	-0.0401	-	7470A	PPB	19 Oct 2020 11:26:10	U	SMPL
PB132434TB - 1	PB132399TB	-7	-0.0134	-	7470A	PPB	19 Oct 2020 11:28:27	U	SMPL
CCV51 - 1	CCV51	17980	4.7356	-	7470A	PPB	19 Oct 2020 11:30:44	U	SMPL
CCB51 - 1	CCB51	-65	-0.0287	-	7470A	PPB	19 Oct 2020 11:33:00	U	SMPL
L4392-18LX5 - 1		24	-0.0052	-	7470A	PPB	19 Oct 2020 11:35:15	U	SMPL
L4392-18A - 1		14721	3.8751	-	7470A	PPB	19 Oct 2020 11:37:31	U	SMPL
L4218-04LX5 - 1		-53	-0.0256	-	7470A	PPB	19 Oct 2020 11:39:46	U	SMPL
L4218-04A - 1		14126	3.7180	-	7470A	PPB	19 Oct 2020 11:42:02	U	SMPL
L4412-06LX5 - 1		-31	-0.0198	-	7470A	PPB	19 Oct 2020 11:44:17	U	SMPL
L4412-06A - 1		15840	4.1706	-	7470A	PPB	19 Oct 2020 11:46:33	U	SMPL
CCV52 - 1	CCV52	18084	4.7630	-	7470A	PPB	19 Oct 2020 11:48:48	U	SMPL
CCB52 - 1	CCB52	-104	-0.0390	-	7470A	PPB	19 Oct 2020 11:51:04	U	SMPL