

LA111593


7471B

INSTRUMENT ID: CV2

Linear

μ Abs.:

43203



A= 0.0000e+000

B= 2.3229e-004

C= 3.2941e-002

Rho= 0.9999158

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.040	0.040	32	0.000	32				
0.2	0.200	0.229	0.029	842	0.0 %	842				
2.5	2.500	2.464	-0.036	10466	0.0 %	10466				
5.0	5.000	4.942	-0.058	21131	0.0 %	21131				
7.5	7.500	7.457	-0.043	31958	0.0 %	31958				
10.0	10.000	10.069	0.069	43203	0.0 %	43203				

LB111593 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Oct 2020 14:46:46

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	32	-	0.0000	7471B	PPB	19 Oct 2020 14:48:08	S	Std
0.2 - 1	50.2	842	-	0.2000	7471B	PPB	19 Oct 2020 14:50:25	S	Std
2.5 - 1	52.5	10466	-	2.5000	7471B	PPB	19 Oct 2020 14:52:43	S	Std
5.0 - 1	55	21131	-	5.0000	7471B	PPB	19 Oct 2020 14:55:00	S	Std
7.5 - 1	57.5	31958	-	7.5000	7471B	PPB	19 Oct 2020 14:57:16	S	Std
10.0 - 1	511	43203	-	10.0000	7471B	PPB	19 Oct 2020 14:59:33	S	Std
ICV63 - 1	ICV63	16599	3.8888	-	7471B	PPB	19 Oct 2020 15:02:19	U	SMPL
ICB63 - 1	ICB63	-77	0.0151	-	7471B	PPB	19 Oct 2020 15:04:34	U	SMPL
CCV53 - 1	CCV53	20731	4.8486	-	7471B	PPB	19 Oct 2020 15:06:49	U	SMPL
CCB53 - 1	CCB53	-80	0.0144	-	7471B	PPB	19 Oct 2020 15:09:04	U	SMPL
CRA - 1	CRA	826	0.2248	-	7471B	PPB	19 Oct 2020 15:11:20	U	SMPL
HighStd - 1	HighStd	42743	9.9619	-	7471B	PPB	19 Oct 2020 15:13:36	U	SMPL
ChkStd - 1	ChkStd	29017	6.7734	-	7471B	PPB	19 Oct 2020 15:15:51	U	SMPL
PB132456BL - 1	PB132456BL	-147	-0.0012	-	7471B	PPB	19 Oct 2020 15:18:09	U	SMPL
PB132456BS - 1	PB132456BS	14690	3.4454	-	7471B	PPB	19 Oct 2020 15:20:26	U	SMPL
L4433-02 - 1	PR-SOIL-PILE	1776	0.4455	-	7471B	PPB	19 Oct 2020 15:22:42	U	SMPL
L4434-01 - 1	40978-S	4202	1.0090	-	7471B	PPB	19 Oct 2020 15:24:59	U	SMPL
L4434-02 - 1	16757-S	6147	1.4609	-	7471B	PPB	19 Oct 2020 15:27:16	U	SMPL
L4435-02 - 1	5412	1959	0.4880	-	7471B	PPB	19 Oct 2020 15:29:33	U	SMPL
CCV54 - 1	CCV54	21448	5.0152	-	7471B	PPB	19 Oct 2020 15:31:48	U	SMPL
CCB54 - 1	CCB54	-39	0.0239	-	7471B	PPB	19 Oct 2020 15:34:04	U	SMPL
L4454-01 - 1	TEST-PIT	5669	1.3498	-	7471B	PPB	19 Oct 2020 15:41:08	U	SMPL
L4454-01DUP - 1	TEST-PITDUP	5198	1.2404	-	7471B	PPB	19 Oct 2020 15:43:22	U	SMPL
L4454-01MS - 1	TEST-PITMS	20169	4.7181	-	7471B	PPB	19 Oct 2020 15:45:38	U	SMPL
L4454-01MSD - 1	TEST-PITMSD	20612	4.8210	-	7471B	PPB	19 Oct 2020 15:47:53	U	SMPL
L4454-01LX5 - 1		913	0.2450	-	7471B	PPB	19 Oct 2020 15:50:09	U	SMPL
L4454-01A - 1		21126	4.9404	-	7471B	PPB	19 Oct 2020 15:52:26	U	SMPL
CCV55 - 1	CCV55	21236	4.9660	-	7471B	PPB	19 Oct 2020 15:54:42	U	SMPL
CCB55 - 1	CCB55	-64	0.0181	-	7471B	PPB	19 Oct 2020 15:57:00	U	SMPL