



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111594

Slope : 98.1

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/19/2020	14:40
2	CAL2	1	Water	NA	NA	20.1	7.00	10/19/2020	14:41
3	CAL3	1	Water	NA	NA	20.1	10.03	10/19/2020	14:42
4	ICV	1	Water	NA	NA	20.1	6.99	10/19/2020	14:45
5	CCV1	1	Water	NA	NA	20.2	2.01	10/19/2020	14:48
6	L4462-01	1	Water	NA	NA	20.2	5.85	10/19/2020	14:50
7	L4462-01DUP	1	Water	NA	NA	20.3	5.86	10/19/2020	14:51
8	CCV2	1	Water	NA	NA	20.3	12.02	10/19/2020	15:00

LB 111594

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4462 WorkList ID : 143459 Department : Wet-Chemistry Date : 10-19-2020 14:34:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/13/2020	Water	L4462-01	-pH	Cool 4 deg C	STER10	P21	POOL-WATER	10/13/2020	9040C

Date/Time 10/19/2020 14:35
Received by: JH / wet-chem
Relinquished by: AM (wet-chem)

Date/Time 10/19/2020 15:35
Received by: AM (wet-chem)
Relinquished by: JH / wet-chem
Chem