



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB111598

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/19/2020	09:45
2	L4434-03		1	9	100.00	>212.0	>212.0	10/19/2020	11:15
3	L4434-03DUP		1	9	100.00	>212.0	>212.0	10/19/2020	11:40
4	L4434-04		1	8	100.00	>212.0	>212.0	10/19/2020	12:05
5	L4434-05		1	9	100.00	>212.0	>212.0	10/19/2020	12:35
6	L4434-06		1	8	100.00	>212.0	>212.0	10/19/2020	12:58
7	L4457-02		1	7	100.00	>212.0	>212.0	10/19/2020	13:25
8	L4457-03		1	8	100.00	>212.0	>212.0	10/19/2020	13:55
9	L4457-04		1	9	100.00	>212.0	>212.0	10/19/2020	14:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 111598

WorkList Name : fp101920 WorkList ID : 143437 Department : Wet-Chemistry Date : 10-19-2020 09:17:18

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/23/2020	Water	L4434-03	Flash Point	Cool 4 deg C	PSEG01	O21	40978-A	10/16/2020	1010A
10/23/2020	Water	L4434-04	Flash Point	Cool 4 deg C	PSEG01	O21	16757-A	10/16/2020	1010A
10/23/2020	Water	L4434-05	Flash Point	Cool 4 deg C	PSEG01	O21	16743	10/16/2020	1010A
10/23/2020	Water	L4434-06	Flash Point	Cool 4 deg C	PSEG01	O21	16804	10/16/2020	1010A
10/31/2020	Water	L4457-02	Flash Point	Cool 4 deg C	DIST01	P22	SUTTON-PR-2152	10/17/2020	1010A
10/31/2020	Water	L4457-03	Flash Point	Cool 4 deg C	DIST01	P22	NOTT-AVE-PR-2150	10/17/2020	1010A
10/31/2020	Water	L4457-04	Flash Point	Cool 4 deg C	DIST01	P22	ROOSEVELT-PR-2151	10/17/2020	1010A

Date/Time 10/19/2020 09:25
 Received by: KS (Wetchem)
 Relinquished by: 99 (Wetchem)

Date/Time 10/19/2020 14:30
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)