



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111602

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/19/2020	16:45
2	CAL2	1	Water	NA	NA	20.1	7.01	10/19/2020	16:46
3	CAL3	1	Water	NA	NA	20.1	10.03	10/19/2020	16:47
4	ICV	1	Water	NA	NA	20.1	6.98	10/19/2020	16:48
5	CCV1	1	Water	NA	NA	20.2	2.01	10/19/2020	16:49
6	L4222-05	1	Solid	20.03	20	20.3	8.21	10/19/2020	16:50
7	L4222-05DUP	1	Solid	20.04	20	20.4	8.22	10/19/2020	16:51
8	L4457-01	1	Solid	20.01	20	18.9	8.45	10/19/2020	17:00
9	L4460-01	1	Solid	20.03	20	20.5	5.83	10/19/2020	17:02
10	CCV2	1	Water	NA	NA	20.3	12.02	10/19/2020	17:05

287 111602

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COROSIVITY L4457 WorkList ID : 143456 Department : Wet-Chemistry Date : 10-19-2020 13:57:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/19/2020	Solid	L4222-05	Corrosivity	Cool 4 deg C	PSEG05	O31	EO-03-101920	10/19/2020	9045D
10/17/2020	Solid	L4457-01	Corrosivity	Cool 4 deg C	DIST01	P22	SUTTON-PR-2152	10/17/2020	9045D
10/19/2020	Solid	L4460-01	Corrosivity	Cool 4 deg C	PSEG01	O31	CAPEMAY-ST-DRUM	10/19/2020	9045D

Date/Time 10-19-2020 16:10
 Received by: JS (Wetchem)
 Relinquished by: JS (Wetchem)

Date/Time 10-19-2020 17:35
 Received by: JS (Wetchem)
 Relinquished by: JS (Wetchem)