



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111607

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/20/2020	11:40
2	CAL2	1	Water	NA	NA	20.1	7.00	10/20/2020	11:41
3	CAL3	1	Water	NA	NA	20.1	10.03	10/20/2020	11:42
4	ICV	1	Water	NA	NA	20.2	7.02	10/20/2020	11:44
5	CCV1	1	Water	NA	NA	20.1	2.02	10/20/2020	11:45
6	L4473-02	1	Water	NA	NA	14.7	6.61	10/20/2020	11:46
7	L4473-02DUP	1	Water	NA	NA	14.8	6.62	10/20/2020	11:47
8	L4473-03	1	Water	NA	NA	15.8	7.36	10/20/2020	11:48
9	L4473-04	1	Water	NA	NA	16.6	6.88	10/20/2020	11:50
10	CCV2	1	Water	NA	NA	20.2	12.02	10/20/2020	11:52

111607

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph 14473 WorkList ID : 143492 Department : Wet-Chemistry Date : 10-20-2020 11:34:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/19/2020	Water	L4473-02 -H	pH	Cool 4 deg C	TETR06	D11	TT-019-IDWGW-20201019	10/19/2020	9040C
10/19/2020	Water	L4473-03 -H	pH	Cool 4 deg C	TETR06	D11	TT-020-IDWGW-20201019	10/19/2020	9040C
10/19/2020	Water	L4473-04 -H	pH	Cool 4 deg C	TETR06	D11	TT-021-IDWGW-20201019	10/19/2020	9040C

Date/Time 10.20.2020 11:35
 Received by: 77 (wet chem)
 Relinquished by: 70

Date/Time 10.20.2020 12:40
 Received by: [Signature]
 Relinquished by: [Signature]