


LB111609

245.1

INSTRUMENT ID: CV2



A= 0.0000e+000

B= 3.5070e-004

C= -9.4293e-002

Rho= 0.9997290

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.093	-0.093	5	0.000	5				
0.05	0.050	0.047	-0.003	403	0.0 %	403	0			
0.2	0.200	0.236	0.036	943	0.0 %	943				
2.5	2.500	2.587	0.087	7646	0.0 %	7646				
5.0	5.000	5.080	0.080	14753	0.0 %	14753				
7.5	7.500	7.340	-0.160	21198	0.0 %	21198				
10.0	10.000	10.053	0.053	28934	0.0 %	28934				

LB111609 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 20 Oct 2020 13:49:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	5	-	0.0000	245.1	PPB	20 Oct 2020 14:18:31	S	Std
0.05 - 1	50.05	403	-	0.0500	245.1	PPB	20 Oct 2020 14:20:48	S	Std
0.2 - 1	50.2	943	-	0.2000	245.1	PPB	20 Oct 2020 14:25:56	S	Std
2.5 - 1	52.5	7646	-	2.5000	245.1	PPB	20 Oct 2020 14:28:13	S	Std
5.0 - 1	55	14753	-	5.0000	245.1	PPB	20 Oct 2020 14:30:29	S	Std
7.5 - 1	57.5	21198	-	7.5000	245.1	PPB	20 Oct 2020 14:32:45	S	Std
10.0 - 1	510	28934	-	10.0000	245.1	PPB	20 Oct 2020 14:37:38	S	Std
ICV65 - 1	ICV65	11263	3.8556	-245.1		PPB	20 Oct 2020 14:40:23	U	SMPL
ICB65 - 1	ICB65	223	-0.0161	-245.1		PPB	20 Oct 2020 14:45:11	U	SMPL
CCV60 - 1	CCV60	15097	5.2002	-245.1		PPB	20 Oct 2020 14:47:26	U	SMPL
CCB60 - 1	CCB60	236	-0.0115	-245.1		PPB	20 Oct 2020 14:49:41	U	SMPL
CRA - 1	CRA	406	0.0481	-245.1		PPB	20 Oct 2020 14:51:56	U	SMPL
HighStd - 1	HighStd	29277	10.1730	-245.1		PPB	20 Oct 2020 14:54:11	U	SMPL
ChkStd - 1	ChkStd	20561	7.1163	-245.1		PPB	20 Oct 2020 14:56:26	U	SMPL
PB132500BL - 1	PBW	276	0.0025	-245.1		PPB	20 Oct 2020 14:58:44	U	SMPL
PB132500BS - 1	LCSW	10566	3.6112	-245.1		PPB	20 Oct 2020 15:01:01	U	SMPL
L4373-02 - 1	EFFLUENT-METALS	1706	0.5040	-245.1		PPB	20 Oct 2020 15:03:18	U	SMPL
L4455-03 - 1	HP-627	285	0.0057	-245.1		PPB	20 Oct 2020 15:05:34	U	SMPL
L4455-03DUP - 1	HP-627DUP	331	0.0218	-245.1		PPB	20 Oct 2020 15:07:51	U	SMPL
L4455-03MS - 1	HP-627MS	10118	3.4540	-245.1		PPB	20 Oct 2020 15:10:08	U	SMPL
L4455-03MSD - 1	HP-627MSD	11710	4.0123	-245.1		PPB	20 Oct 2020 15:12:24	U	SMPL
CCV61 - 1	CCV61	14351	4.9385	-245.1		PPB	20 Oct 2020 15:14:39	U	SMPL
CCB61 - 1	CCB61	290	0.0074	-245.1		PPB	20 Oct 2020 15:19:37	U	SMPL
L4456-03 - 1	HP-640	92	-0.0620	-245.1		PPB	20 Oct 2020 15:21:52	U	SMPL
L4457-02 - 1	SUTTON-PR-2152	5734	1.9166	-245.1		PPB	20 Oct 2020 15:24:07	U	SMPL
L4457-03 - 1	NOTT-AVE-PR-2150	154	-0.0403	-245.1		PPB	20 Oct 2020 15:26:22	U	SMPL
L4457-04 - 1	ROOSEVELT-PR-2151	280	0.0039	-245.1		PPB	20 Oct 2020 15:28:39	U	SMPL
L4455-03LX5 - 1		383	0.0400	-245.1		PPB	20 Oct 2020 15:30:55	U	SMPL
L4455-03A - 1		10862	3.7150	-245.1		PPB	20 Oct 2020 15:33:11	U	SMPL
CCV62 - 1	CCV62	14228	4.8954	-245.1		PPB	20 Oct 2020 15:35:28	U	SMPL
CCB62 - 1	CCB62	307	0.0134	-245.1		PPB	20 Oct 2020 15:40:19	U	SMPL