



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111611

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

True Value of CCV3 = 2.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/20/2020	12:00
2	CAL2	1	Water	NA	NA	20.2	7.00	10/20/2020	12:01
3	CAL3	1	Water	NA	NA	20.2	10.03	10/20/2020	12:02
4	ICV	1	Water	NA	NA	20.2	6.98	10/20/2020	12:05
5	CCV1	1	Water	NA	NA	20.1	2.01	10/20/2020	12:06
6	L4436-12	1	Solid	20.03	20	21.2	10.85	10/20/2020	12:08
7	L4436-12DUP	1	Solid	20.04	20	21.3	10.86	10/20/2020	12:09
8	L4436-13	1	Solid	20.01	20	21.1	9.01	10/20/2020	12:10
9	L4436-14	1	Solid	20.03	20	21.0	8.64	10/20/2020	12:11
10	L4436-15	1	Solid	20.04	20	21.2	8.91	10/20/2020	12:12
11	L4436-16	1	Solid	20.05	20	21.3	8.58	10/20/2020	12:15
12	L4436-17	1	Solid	20.04	20	21.1	8.88	10/20/2020	12:18
13	L4436-18	1	Solid	20.02	20	21.0	8.87	10/20/2020	12:20
14	L4436-19	1	Solid	20.03	20	20.8	8.67	10/20/2020	12:21
15	L4436-20	1	Solid	20.04	20	21.3	8.49	10/20/2020	12:22
16	CCV2	1	Water	NA	NA	20.2	12.03	10/20/2020	12:23
17	L4436-21	1	Solid	20.04	20	21.0	8.74	10/20/2020	12:25
18	CCV3	1	Water	NA	NA	20.2	2.02	10/20/2020	12:26

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L4436

WorkList ID : 143495

Department : Wet-Chemistry

Date : 10-20-2020 11:38:08

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/19/2020	Solid	L4436-12	F	Corrosivity	JPCL01	P11	33A	10/19/2020	9045D
10/19/2020	Solid	L4436-13	F	Corrosivity	JPCL01	P11	33B	10/19/2020	9045D
10/19/2020	Solid	L4436-14	F	Corrosivity	JPCL01	P11	34A	10/19/2020	9045D
10/19/2020	Solid	L4436-15	F	Corrosivity	JPCL01	P11	34B	10/19/2020	9045D
10/19/2020	Solid	L4436-16	F	Corrosivity	JPCL01	P11	35A	10/19/2020	9045D
10/19/2020	Solid	L4436-17	F	Corrosivity	JPCL01	P11	35B	10/19/2020	9045D
10/19/2020	Solid	L4436-18	F	Corrosivity	JPCL01	P11	36A	10/19/2020	9045D
10/19/2020	Solid	L4436-19	F	Corrosivity	JPCL01	P11	36B	10/19/2020	9045D
10/19/2020	Solid	L4436-20	F	Corrosivity	JPCL01	P11	37A	10/19/2020	9045D
10/19/2020	Solid	L4436-21	F	Corrosivity	JPCL01	P11	37B	10/19/2020	9045D

Date/Time 10-20-2020 11:40
 Received by: Jo [Signature]
 Relinquished by: [Signature]

Date/Time 10-20-2020 14:10
 Received by: [Signature]
 Relinquished by: [Signature]