



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 10/21/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 16:35
In Date: 10/20/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 10/21/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB111613

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L4221-10	OK-03-102020	30	1.13	9.53	8.89	92.4	
L4221-11	OK-03-102020	31	1.14	9.96	9.28	92.3	
L4221-12	OK-03-102020-1	32	1.11	9.57	8.93	92.4	
L4221-13	OK-03-102020-2	33	1.13	9.78	8.81	88.8	
L4221-13D	OK-03-102020-2DUP	34	1.17	9.79	8.84	89.0	
L4297-06	RO-RBR-200031	1	1.13	9.62	3.39	26.6	GEL MATRIX
L4297-07	RO-RBR-200034	2	1.15	9.98	1.98	9.4	GEL MATRIX
L4436-02	33A	3	1.12	9.62	8.31	84.6	
L4436-03	33B	4	1.13	9.56	8.21	84.0	
L4436-04	34A	5	1.13	9.83	8.5	84.7	
L4436-05	34B	6	1.15	9.94	9.04	89.8	
L4436-06	35A	7	1.16	9.94	9.14	90.9	
L4436-07	35B	8	1.13	9.9	8.92	88.8	
L4436-08	36A	9	1.13	9.71	8.97	91.4	
L4436-09	36B	10	1.15	9.93	8.94	88.7	
L4436-10	37A	11	1.14	9.96	9.12	90.5	
L4436-11	37B	12	1.13	9.82	8.54	85.3	
L4463-01	VNJ-238	13	1.15	9.95	9.00	89.2	
L4463-02	VNJ-238	14	1.13	9.77	8.73	88.0	
L4467-01	VB-15242	26	1.13	9.82	7.3	71.0	
L4467-03	VNJ-240	27	1.14	9.94	8.38	82.3	
L4467-05	357	28	1.13	9.63	8.1	82.0	
L4468-01	72-11977	29	1.17	9.97	9.12	90.3	
L4472-01	PCB-101720-SR-01	15	1.00	2.00	2.00	100.0	CAULKING SAMPLE
L4472-02	PCB-101720-SR-02	16	1.00	2.00	2.00	100.0	CAULKING SAMPLE
L4472-03	PCB-101720-SR-03	17	1.00	2.00	2.00	100.0	CAULKING SAMPLE
L4472-04	PCB-101720-SR-04	18	1.00	2.00	2.00	100.0	CAULKING SAMPLE
L4475-01	C-1-0.5-1.0	19	1.13	9.63	8.73	89.4	



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 10/21/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 16:35
In Date: 10/20/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 10/21/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB111613

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
L4475-02	C-2-0.5-1.0	20	1.13	9.96	8.2	80.1	
L4476-01	PE-41-(6.5-7.0)	21	1.14	9.98	8.75	86.1	
L4476-02	PE-41-(8.0-8.5)	22	1.14	9.88	8.62	85.6	
L4476-03	PE-42-(6.5-7.0)	23	1.1	9.74	8.6	86.8	
L4476-04	PE-42-(8.0-8.5)	24	1.14	9.96	9.00	89.1	
L4476-05	PE-47-(8.0-8.5)	25	1.18	9.98	9.17	90.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$

111613

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-102020 WorkList ID : 143485 Department : Wet-Chemistry Date : 10-20-2020 09:14:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/23/2020	Solid	L4221-10	Percent Solids	Cool 4 deg C	PSEG05		OK-03-102020	10/20/2020	Chemtech -SO
10/23/2020	Solid	L4221-11	Percent Solids	Cool 4 deg C	PSEG05		OK-03-102020	10/20/2020	Chemtech -SO
10/23/2020	Solid	L4221-12	Percent Solids	Cool 4 deg C	PSEG05		OK-03-102020-1	10/20/2020	Chemtech -SO
10/23/2020	Solid	L4221-13	Percent Solids	Cool 4 deg C	PSEG05		OK-03-102020-2	10/20/2020	Chemtech -SO
10/14/2020	Solid	L4297-06	Percent Solids	Cool 4 deg C	PSEG01	O41	RO-RBR-200031	10/07/2020	Chemtech -SO
10/14/2020	Solid	L4297-07	Percent Solids	Cool 4 deg C	PSEG01	O41	RO-RBR-200034	10/07/2020	Chemtech -SO
10/29/2020	Solid	L4436-02	Percent Solids	Cool 4 deg C	JPCL01	P11	33A	10/19/2020	Chemtech -SO
10/29/2020	Solid	L4436-03	Percent Solids	Cool 4 deg C	JPCL01	P11	33B	10/19/2020	Chemtech -SO
10/29/2020	Solid	L4436-04	Percent Solids	Cool 4 deg C	JPCL01	P11	34A	10/19/2020	Chemtech -SO
10/29/2020	Solid	L4436-05	Percent Solids	Cool 4 deg C	JPCL01	P11	34B	10/19/2020	Chemtech -SO
10/29/2020	Solid	L4436-06	Percent Solids	Cool 4 deg C	JPCL01	P11	35A	10/19/2020	Chemtech -SO
10/29/2020	Solid	L4436-07	Percent Solids	Cool 4 deg C	JPCL01	P11	35B	10/19/2020	Chemtech -SO
10/29/2020	Solid	L4436-08	Percent Solids	Cool 4 deg C	JPCL01	P11	36A	10/19/2020	Chemtech -SO
10/29/2020	Solid	L4436-09	Percent Solids	Cool 4 deg C	JPCL01	P11	36B	10/19/2020	Chemtech -SO
10/29/2020	Solid	L4436-10	Percent Solids	Cool 4 deg C	JPCL01	P11	37A	10/19/2020	Chemtech -SO
10/29/2020	Solid	L4436-11	Percent Solids	Cool 4 deg C	JPCL01	P11	37B	10/19/2020	Chemtech -SO
10/26/2020	Solid	L4463-01	Percent Solids	Cool 4 deg C	PSEG01	O31	VNJ-238	10/19/2020	Chemtech -SO
10/26/2020	Solid	L4463-02	Percent Solids	Cool 4 deg C	PSEG01	O31	VNJ-238	10/19/2020	Chemtech -SO
10/23/2020	Solid	L4467-01	Percent Solids	Cool 4 deg C	PSEG01		VB-15242	10/20/2020	Chemtech -SO
10/23/2020	Solid	L4467-03	Percent Solids	Cool 4 deg C	PSEG01		VNJ-240	10/20/2020	Chemtech -SO
10/23/2020	Solid	L4467-05	Percent Solids	Cool 4 deg C	PSEG01		357	10/20/2020	Chemtech -SO
10/23/2020	Solid	L4468-01	Percent Solids	Cool 4 deg C	PSEG01	D12	72-11977	10/20/2020	Chemtech -SO

Date/Time 10-20-2020 17:30
 Received by: JP (Wet Chem)
 Relinquished by: ad (sm)

Date/Time 10-20-2020 17:30
 Received by: CP (sm)
 Relinquished by: JP (Wet Chem)

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-102020 WorkList ID : 143485 Department : Wet-Chemistry Date : 10-20-2020 09:14:54

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/27/2020	Solid	L4472-01	Percent Solids	Cool 4 deg C	CLOU03	D11	PCB-101720-SR-01	10/17/2020	Chemtech -SO
10/27/2020	Solid	L4472-02	Percent Solids	Cool 4 deg C	CLOU03	D11	PCB-101720-SR-02	10/17/2020	Chemtech -SO
10/27/2020	Solid	L4472-03	Percent Solids	Cool 4 deg C	CLOU03	D11	PCB-101720-SR-03	10/17/2020	Chemtech -SO
10/27/2020	Solid	L4472-04	Percent Solids	Cool 4 deg C	CLOU03	D11	PCB-101720-SR-04	10/17/2020	Chemtech -SO
10/24/2020	Solid	L4475-01	Percent Solids	Cool 4 deg C	POWE02	D11	C-1-0.5-1.0	10/14/2020	Chemtech -SO
10/24/2020	Solid	L4475-02	Percent Solids	Cool 4 deg C	POWE02	D11	C-2-0.5-1.0	10/14/2020	Chemtech -SO
10/21/2020	Solid	L4476-01	Percent Solids	Cool 4 deg C	PSEG01	D12	PE-41-(6.5-7.0)	10/20/2020	Chemtech -SO
10/21/2020	Solid	L4476-02	Percent Solids	Cool 4 deg C	PSEG01	D12	PE-41-(8.0-8.5)	10/20/2020	Chemtech -SO
10/21/2020	Solid	L4476-03	Percent Solids	Cool 4 deg C	PSEG01	D12	PE-42-(6.5-7.0)	10/20/2020	Chemtech -SO
10/21/2020	Solid	L4476-04	Percent Solids	Cool 4 deg C	PSEG01	D12	PE-42-(8.0-8.5)	10/20/2020	Chemtech -SO
10/21/2020	Solid	L4476-05	Percent Solids	Cool 4 deg C	PSEG01	D12	PE-47-(8.0-8.5)	10/20/2020	Chemtech -SO

Date/Time 10/20/2020 16:20
 Received by: JP West Chem
 Relinquished by: CP SMI

Date/Time 10/20/2020 17:30
 Received by: CP SMI
 Relinquished by: JP West Chem