

LB111614


7471B

INSTRUMENT ID: CV2

Linear

 μ Abs.:

41820



A= 0.0000e+000

B= 2.4381e-004

C= -1.9598e-001

Rho= 0.9998413

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.050	0.050	1008	0.000	1008				
0.2	0.200	0.237	0.037	1775	0.0 %	1775				
2.5	2.500	2.357	-0.143	10473	0.0 %	10473				
5.0	5.000	5.039	0.039	21471	0.0 %	21471				
7.5	7.500	7.517	0.017	31633	0.0 %	31633				
10.0	10.000	10.000	0.000	41820	0.0 %	41820				

LB111614 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 20 Oct 2020 15:54:03

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	1008	-	0.0000	7471B	PPB	20 Oct 2020 15:55:00	S	Std
0.2 - 1	50.2	1775	-	0.2000	7471B	PPB	20 Oct 2020 15:57:17	S	Std
2.5 - 1	52.1	10473	-	2.5000	7471B	PPB	20 Oct 2020 15:59:34	S	Std
5.0 - 1	55	21471	-	5.0000	7471B	PPB	20 Oct 2020 16:01:51	S	Std
7.5 - 1	57.1	31633	-	7.5000	7471B	PPB	20 Oct 2020 16:04:08	S	Std
10.0 - 1	58	41820	-	10.0000	7471B	PPB	20 Oct 2020 16:09:00	S	Std
ICV66 - 1	ICV66	17003	3.9496	-	7471B	PPB	20 Oct 2020 16:11:43	U	SMPL
ICB66 - 1	ICB66	666	-0.0336	-	7471B	PPB	20 Oct 2020 16:13:58	U	SMPL
CCV63 - 1	CCV63	21557	5.0599	-	7471B	PPB	20 Oct 2020 16:16:14	U	SMPL
CCB63 - 1	CCB63	666	-0.0336	-	7471B	PPB	20 Oct 2020 16:18:29	U	SMPL
CRA - 1	CRA	1727	0.2251	-	7471B	PPB	20 Oct 2020 16:20:45	U	SMPL
HighStd - 1	HighStd	42224	10.0989	-	7471B	PPB	20 Oct 2020 16:22:59	U	SMPL
ChkStd - 1	ChkStd	29156	6.9127	-	7471B	PPB	20 Oct 2020 16:25:14	U	SMPL
PB132501BL - 1	PBS	769	-0.0085	-	7471B	PPB	20 Oct 2020 16:30:04	U	SMPL
PB132501BS - 1	LCSS	16675	3.8696	-	7471B	PPB	20 Oct 2020 16:34:51	U	SMPL
L4222-01 - 1	EO-03-101920	5150	1.0597	-	7471B	PPB	20 Oct 2020 16:37:07	U	SMPL
L4222-01DUP - 1	EO-03-101920DUP	5271	1.0892	-	7471B	PPB	20 Oct 2020 16:39:25	U	SMPL
L4222-01MS - 1	EO-03-101920MS	18992	4.4345	-	7471B	PPB	20 Oct 2020 16:41:42	U	SMPL
L4222-01MSD - 1	EO-03-101920MSD	20200	4.7291	-	7471B	PPB	20 Oct 2020 16:43:58	U	SMPL
L4458-02 - 1	BS-DRUM	7482	1.6282	-	7471B	PPB	20 Oct 2020 16:46:14	U	SMPL
CCV64 - 1	CCV64	21255	4.9863	-	7471B	PPB	20 Oct 2020 16:48:30	U	SMPL
CCB64 - 1	CCB64	725	-0.0192	-	7471B	PPB	20 Oct 2020 16:50:46	U	SMPL
L4460-01 - 1	CAPEMAY-ST-DRUM	88	-0.1745	-	7471B	PPB	20 Oct 2020 16:53:02	U	SMPL
L4461-01 - 1	WCS-OH-F1-5.0-10.0	11904	2.7064	-	7471B	PPB	20 Oct 2020 16:55:17	U	SMPL
L4461-02 - 1	WCS-OH-B2-0.2-4.5	96375	23.3017	-	7471B	PPB	20 Oct 2020 16:57:33	U	SMPL
L4222-01LX5 - 1		1310	0.1234	-	7471B	PPB	20 Oct 2020 16:59:49	U	SMPL
L4222-01A - 1		19940	4.6657	-	7471B	PPB	20 Oct 2020 17:02:07	U	SMPL
L4461-02DLX5 - 1		20144	4.7154	-	7471B	PPB	20 Oct 2020 17:04:23	U	SMPL
CCV65 - 1	CCV65	20324	4.7593	-	7471B	PPB	20 Oct 2020 17:06:40	U	SMPL
CCB65 - 1	CCB65	675	-0.0314	-	7471B	PPB	20 Oct 2020 17:08:57	U	SMPL