


LB111623

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 3.2948e-004

C= -3.7731e-002

Rho= 0.9995772

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.026	0.026	193	0.000	193				
0.2	0.200	0.208	0.008	746	0.0 %	746				
2.5	2.500	2.344	-0.156	7229	0.0 %	7229				
5.0	5.000	5.118	0.118	15647	0.0 %	15647				
7.5	7.500	7.625	0.125	23257	0.0 %	23257				
10.0	10.000	9.879	-0.121	30099	0.0 %	30099				

LB111623 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 21 Oct 2020 09:20:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	50	193	-	0.00007470A	PPB	21 Oct 2020 09:25:19	S	Std
0.2 - 1	50.2	746	-	0.20007470A	PPB	21 Oct 2020 09:27:36	S	Std
2.5 - 1	52.5	7229	-	2.50007470A	PPB	21 Oct 2020 09:29:53	S	Std
5.0 - 1	55	15647	-	5.00007470A	PPB	21 Oct 2020 09:32:10	S	Std
7.5 - 1	57.5	23257	-	7.50007470A	PPB	21 Oct 2020 09:34:26	S	Std
10.0 - 1	510	30099	-	10.00007470A	PPB	21 Oct 2020 09:39:18	S	Std
ICV67 - 1	ICV67	12364	4.0360	-7470A	PPB	21 Oct 2020 09:41:56	U	SMPL
ICB67 - 1	ICB67	-9	-0.0407	-7470A	PPB	21 Oct 2020 09:44:10	U	SMPL
CCV66 - 1	CCV66	15561	5.0893	-7470A	PPB	21 Oct 2020 09:46:25	U	SMPL
CCB66 - 1	CCB66	59	-0.0183	-7470A	PPB	21 Oct 2020 09:48:40	U	SMPL
CRA - 1	CRA	685	0.1880	-7470A	PPB	21 Oct 2020 09:50:56	U	SMPL
HighStd - 1	HighStd	29466	9.6708	-7470A	PPB	21 Oct 2020 09:53:10	U	SMPL
ChkStd - 1	ChkStd	21504	7.0474	-7470A	PPB	21 Oct 2020 09:55:25	U	SMPL
PB132515BL - 1	PBW	17	-0.0321	-7470A	PPB	21 Oct 2020 10:02:31	U	SMPL
PB132515BS - 1	LCSW	12085	3.9440	-7470A	PPB	21 Oct 2020 10:04:46	U	SMPL
L4222-05 - 1	EO-03-101920	581	0.1537	-7470A	PPB	21 Oct 2020 10:07:03	U	SMPL
L4222-05DUP - 1	EO-03-101920DUP	644	0.1745	-7470A	PPB	21 Oct 2020 10:11:58	U	SMPL
L4222-05MS - 1	EO-03-101920MS	11750	3.8337	-7470A	PPB	21 Oct 2020 10:17:03	U	SMPL
L4222-05MSD - 1	EO-03-101920MSD	11626	3.7928	-7470A	PPB	21 Oct 2020 10:19:20	U	SMPL
L4457-01 - 1	SUTTON-PR-2152-SED	673	0.1840	-7470A	PPB	21 Oct 2020 10:24:55	U	SMPL
CCV67 - 1	CCV67	14274	4.6653	-7470A	PPB	21 Oct 2020 10:27:09	U	SMPL
CCB67 - 1	CCB67	53	-0.0203	-7470A	PPB	21 Oct 2020 10:29:24	U	SMPL
L4459-01 - 1	RBR-20555	608	0.1626	-7470A	PPB	21 Oct 2020 10:31:39	U	SMPL
PB132516BL - 1	PBW	-6	-0.0397	-7470A	PPB	21 Oct 2020 10:36:25	U	SMPL
PB132516BS - 1	LCSW	11320	3.6920	-7470A	PPB	21 Oct 2020 10:38:40	U	SMPL
L4473-02 - 1	TT-019-IDWGW-20201019	711	0.1965	-7470A	PPB	21 Oct 2020 10:40:56	U	SMPL
L4473-02DUP - 1	TT-019-IDWGW-20201019DUP	622	0.1672	-7470A	PPB	21 Oct 2020 10:43:12	U	SMPL
L4473-02MS - 1	TT-019-IDWGW-20201019MS	11505	3.7529	-7470A	PPB	21 Oct 2020 10:45:28	U	SMPL
L4473-02MSD - 1	TT-019-IDWGW-20201019MSD	11788	3.8462	-7470A	PPB	21 Oct 2020 10:47:44	U	SMPL
L4473-03 - 1	TT-020-IDWGW-20201019	289	0.0575	-7470A	PPB	21 Oct 2020 10:50:37	U	SMPL
L4473-04 - 1	TT-021-IDWGW-20201019	584	0.1547	-7470A	PPB	21 Oct 2020 10:52:54	U	SMPL
CCV68 - 1	CCV68	15039	4.9173	-7470A	PPB	21 Oct 2020 11:39:31	U	SMPL
CCB68 - 1	CCB68	1	-0.0374	-7470A	PPB	21 Oct 2020 11:41:46	U	SMPL
PB132515TB - 1	PB132469TB	239	0.0410	-7470A	PPB	21 Oct 2020 11:46:35	U	SMPL
L4222-05LX5 - 1		180	0.0216	-7470A	PPB	21 Oct 2020 11:48:50	U	SMPL
L4222-05A - 1		10614	3.4594	-7470A	PPB	21 Oct 2020 11:51:06	U	SMPL
L4473-02LX5 - 1		153	0.0127	-7470A	PPB	21 Oct 2020 11:53:22	U	SMPL
L4473-02A - 1		10541	3.4353	-7470A	PPB	21 Oct 2020 11:55:37	U	SMPL
CCV69 - 1	CCV69	15698	5.1345	-7470A	PPB	21 Oct 2020 11:57:53	U	SMPL
CCB69 - 1	CCB69	13	-0.0334	-7470A	PPB	21 Oct 2020 12:00:09	U	SMPL