



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111625

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/21/2020	10:40
2	CAL2	1	Water	NA	NA	20.1	7.00	10/21/2020	10:41
3	CAL3	1	Water	NA	NA	20.1	10.03	10/21/2020	10:42
4	ICV	1	Water	NA	NA	20.1	6.98	10/21/2020	10:45
5	CCV1	1	Water	NA	NA	20.2	2.01	10/21/2020	10:46
6	L4457-01	1	Solid	20.02	20	18.9	8.45	10/21/2020	10:49
7	L4457-01DUP	1	Solid	20.03	20	18.8	8.46	10/21/2020	10:50
8	CCV2	1	Water	NA	NA	20.1	12.02	10/21/2020	10:51

WORKLIST(Hardcopy Internal Chain)

78 111625

WorkList Name : PH L4482 WorkList ID : 143524 Department : Wet-Chemistry Date : 10-21-2020 10:27:58

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/17/2020	Solid	L4457-01	pH	Cool 4 deg C	DIST01	P22	SUTTON-PR-2152-SED	10/17/2020	9045D

Date/Time 10-21-2020 10:30
 Received by: 78 [Signature]
 Relinquished by: 78 [Signature]

Date/Time 10-21-2020 11:30
 Received by: 78 [Signature]
 Relinquished by: 78 [Signature]