



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111626

Slope : 97.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/21/2020	10:40
2	CAL2	1	Water	NA	NA	20.1	7.00	10/21/2020	10:41
3	CAL3	1	Water	NA	NA	20.1	10.03	10/21/2020	10:42
4	ICV	1	Water	NA	NA	20.1	6.98	10/21/2020	10:45
5	CCV1	1	Water	NA	NA	20.2	2.01	10/21/2020	10:46
6	L4482-01	1	Water	NA	NA	20.7	4.69	10/21/2020	11:00
7	L4482-01DUP	1	Water	NA	NA	20.8	4.70	10/21/2020	11:01
8	CCV2	1	Water	NA	NA	20.1	12.02	10/21/2020	11:03

18 111626

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4487 WorkList ID : 143525 Department : Wet-Chemistry Date : 10-21-2020 10:29:01

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/19/2020	Water	L4482-01 - 18	pH	Cool 4 deg C	STER14	D21	MINI-CELL	10/19/2020	SM 4500-H B

Date/Time 10-21-2020 11:50
 Received by: AM (wet-chem)
 Relinquished by: 20/601 Chen

Date/Time 10-21-2020 10:30
 Received by: 20/601 - Chen
 Relinquished by: AM (wet-chem)