

LB111646


7471B

INSTRUMENT ID: CV1

Linear

μ Abs.:

36766



A= 0.0000e+000

B= 2.6898e-004

C= -5.0453e-002

Rho= 0.9992620

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.043	-0.043	26	0.000	26				
0.2	0.200	0.192	-0.008	901	0.0 %	901				
2.5	2.500	2.383	-0.117	9046	0.0 %	9046				
5.0	5.000	5.276	0.276	19801	0.0 %	19801				
7.5	7.500	7.554	0.054	28271	0.0 %	28271				
10.0	10.000	9.839	-0.161	36766	0.0 %	36766				

LB111640 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 21 Oct 2020 14:45:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SC	26	-	0.0000	7471B	PPB	21 Oct 2020 14:45:21	S	Std
0.2 - 1	SD-2	901	-	0.2000	7471B	PPB	21 Oct 2020 14:47:39	S	Std
2.5 - 1	SD-5	9046	-	2.5000	7471B	PPB	21 Oct 2020 14:49:56	S	Std
5.0 - 1	SD	19801	-	5.0000	7471B	PPB	21 Oct 2020 14:52:13	S	Std
7.5 - 1	SD-1	28271	-	7.5000	7471B	PPB	21 Oct 2020 14:54:29	S	Std
10.0 - 1	SD	36766	-	10.0000	7471B	PPB	21 Oct 2020 14:56:46	S	Std
ICV68 - 1	ICV68	13680	3.6293	-	7471B	PPB	21 Oct 2020 15:00:16	U	SMPL
ICB68 - 1	ICB68	-78	-0.0714	-	7471B	PPB	21 Oct 2020 15:02:31	U	SMPL
CCV70 - 1	CCV70	19446	5.1802	-	7471B	PPB	21 Oct 2020 15:07:33	U	SMPL
CCB70 - 1	CCB70	-71	-0.0696	-	7471B	PPB	21 Oct 2020 15:09:48	U	SMPL
CRA - 1	CRA	887	0.1881	-	7471B	PPB	21 Oct 2020 15:12:04	U	SMPL
HighStd - 1	HighStd	41415	11.0896	-	7471B	PPB	21 Oct 2020 15:14:19	U	SMPL
ChkStd - 1	ChkStd	30205	8.0742	-	7471B	PPB	21 Oct 2020 15:16:34	U	SMPL
PB132525BL - 1	PBS	-157	-0.0927	-	7471B	PPB	21 Oct 2020 15:18:52	U	SMPL
PB132525BS - 1	LCSS	14637	3.8867	-	7471B	PPB	21 Oct 2020 15:21:09	U	SMPL
L4221-10 - 1	OK-03-102020	1084	0.2411	-	7471B	PPB	21 Oct 2020 15:23:26	U	SMPL
L4436-02 - 1	33A	5579	1.4502	-	7471B	PPB	21 Oct 2020 15:25:43	U	SMPL
L4436-03 - 1	33B	3706	0.9464	-	7471B	PPB	21 Oct 2020 15:27:59	U	SMPL
L4436-04 - 1	34A	3933	1.0075	-	7471B	PPB	21 Oct 2020 15:30:17	U	SMPL
L4436-05 - 1	34B	3440	0.8749	-	7471B	PPB	21 Oct 2020 15:33:13	U	SMPL
CCV71 - 1	CCV71	19661	5.2381	-	7471B	PPB	21 Oct 2020 15:39:26	U	SMPL
CCB71 - 1	CCB71	-79	-0.0717	-	7471B	PPB	21 Oct 2020 15:41:41	U	SMPL
L4436-06 - 1	35A	3964	1.0158	-	7471B	PPB	21 Oct 2020 15:43:57	U	SMPL
L4436-07 - 1	35B	3551	0.9047	-	7471B	PPB	21 Oct 2020 15:46:12	U	SMPL
L4436-08 - 1	36A	1777	0.4275	-	7471B	PPB	21 Oct 2020 15:48:29	U	SMPL
L4436-09 - 1	36B	3235	0.8197	-	7471B	PPB	21 Oct 2020 15:52:28	U	SMPL
L4436-10 - 1	37A	3068	0.7748	-	7471B	PPB	21 Oct 2020 15:54:43	U	SMPL
L4436-11 - 1	37B	12984	3.4420	-	7471B	PPB	21 Oct 2020 15:57:00	U	SMPL
L4463-02 - 1	VNJ-238	5219	1.3534	-	7471B	PPB	21 Oct 2020 15:59:16	U	SMPL
L4467-01 - 1	VB-15242	5922	1.5425	-	7471B	PPB	21 Oct 2020 16:01:34	U	SMPL
L4467-03 - 1	VNJ-240	2051	0.5012	-	7471B	PPB	21 Oct 2020 16:03:51	U	SMPL
L4467-05 - 1	357	4789	1.2377	-	7471B	PPB	21 Oct 2020 16:06:07	U	SMPL
CCV72 - 1	CCV72	19728	5.2561	-	7471B	PPB	21 Oct 2020 16:08:22	U	SMPL
CCB72 - 1	CCB72	-59	-0.0663	-	7471B	PPB	21 Oct 2020 16:10:38	U	SMPL
L4468-01 - 1	72-11977	3175	0.8036	-	7471B	PPB	21 Oct 2020 16:12:54	U	SMPL
L4468-01DUP - 1	72-11977DUP	2983	0.7519	-	7471B	PPB	21 Oct 2020 16:15:10	U	SMPL
L4468-01MS - 1	72-11977MS	20605	5.4920	-	7471B	PPB	21 Oct 2020 16:17:26	U	SMPL
L4468-01MSD - 1	72-11977MSD	18530	4.9338	-	7471B	PPB	21 Oct 2020 16:19:41	U	SMPL
L4468-01LX5 - 1		505	0.0854	-	7471B	PPB	21 Oct 2020 16:21:59	U	SMPL
L4468-01A - 1		19980	5.3239	-	7471B	PPB	21 Oct 2020 16:24:16	U	SMPL
CCV73 - 1	CCV73	19769	5.2671	-	7471B	PPB	21 Oct 2020 16:26:32	U	SMPL
CCB73 - 1	CCB73	-83	-0.0728	-	7471B	PPB	21 Oct 2020 16:28:50	U	SMPL