


LB111649

7470A

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.068	0.068	8	0.000	8				
0.2	0.200	0.226	0.026	561	0.0 %	561				
2.5	2.500	2.323	-0.177	7899	0.0 %	7899				
5.0	5.000	5.174	0.174	17877	0.0 %	17877				
7.5	7.500	7.350	-0.150	25491	0.0 %	25491				
10.0	10.000	10.060	0.060	34976	0.0 %	34976				

LB111649 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 22 Oct 2020 09:16:50

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	8	-	0.0000	7470A	PPB	22 Oct 2020 09:20:02	S	Std
0.2 - 1	50.2	561	-	0.2000	7470A	PPB	22 Oct 2020 09:22:19	S	Std
2.5 - 1	52.5	7899	-	2.5000	7470A	PPB	22 Oct 2020 09:24:36	S	Std
5.0 - 1	55	17877	-	5.0000	7470A	PPB	22 Oct 2020 09:26:53	S	Std
7.5 - 1	57.5	25491	-	7.5000	7470A	PPB	22 Oct 2020 09:29:09	S	Std
10.0 - 1	510	34976	-	10.0000	7470A	PPB	22 Oct 2020 09:34:04	S	Std
ICV69 - 1	ICV69	14973	4.3441	-	7470A	PPB	22 Oct 2020 09:38:02	U	SMPL
ICB69 - 1	ICB69	-126	0.0294	-	7470A	PPB	22 Oct 2020 09:40:17	U	SMPL
CCV74 - 1	CCV74	17705	5.1248	-	7470A	PPB	22 Oct 2020 09:42:31	U	SMPL
CCB74 - 1	CCB74	-146	0.0237	-	7470A	PPB	22 Oct 2020 09:44:46	U	SMPL
CRA - 1	CRA	535	0.2183	-	7470A	PPB	22 Oct 2020 09:47:01	U	SMPL
HighStd - 1	HighStd	35490	10.2071	-	7470A	PPB	22 Oct 2020 09:49:16	U	SMPL
ChkStd - 1	ChkStd	25653	7.3961	-	7470A	PPB	22 Oct 2020 09:51:32	U	SMPL
PB132533BL - 1	PBW	-204	0.0071	-	7470A	PPB	22 Oct 2020 09:53:49	U	SMPL
PB132533BS - 1	LCSW	13230	3.8460	-	7470A	PPB	22 Oct 2020 09:56:06	U	SMPL
L4221-14 - 1	OK-03-102020	-127	0.0291	-	7470A	PPB	22 Oct 2020 09:58:23	U	SMPL
L4221-14DUP - 1	OK-03-102020DUP	-34	0.0557	-	7470A	PPB	22 Oct 2020 10:00:40	U	SMPL
L4221-14MS - 1	OK-03-102020MS	13803	4.0098	-	7470A	PPB	22 Oct 2020 10:02:56	U	SMPL
L4221-14MSD - 1	OK-03-102020MSD	12969	3.7715	-	7470A	PPB	22 Oct 2020 10:05:13	U	SMPL
L4436-12 - 1	33A	-103	0.0360	-	7470A	PPB	22 Oct 2020 10:07:28	U	SMPL
CCV75 - 1	CCV75	17593	5.0928	-	7470A	PPB	22 Oct 2020 10:15:16	U	SMPL
CCB75 - 1	CCB75	-109	0.0343	-	7470A	PPB	22 Oct 2020 10:17:31	U	SMPL
L4436-13 - 1	33B	-13	0.0617	-	7470A	PPB	22 Oct 2020 10:19:46	U	SMPL
L4436-14 - 1	34A	10	0.0683	-	7470A	PPB	22 Oct 2020 10:24:01	U	SMPL
L4436-15 - 1	34B	2	0.0660	-	7470A	PPB	22 Oct 2020 10:26:17	U	SMPL
L4436-16 - 1	35A	-4	0.0643	-	7470A	PPB	22 Oct 2020 10:28:33	U	SMPL
L4436-17 - 1	35B	-15	0.0611	-	7470A	PPB	22 Oct 2020 10:30:48	U	SMPL
L4436-18 - 1	36A	-3	0.0646	-	7470A	PPB	22 Oct 2020 10:33:05	U	SMPL
L4436-19 - 1	36B	18	0.0706	-	7470A	PPB	22 Oct 2020 10:45:49	U	SMPL
L4436-20 - 1	37A	15	0.0697	-	7470A	PPB	22 Oct 2020 10:48:05	U	SMPL
L4436-21 - 1	37B	9	0.0680	-	7470A	PPB	22 Oct 2020 10:50:22	U	SMPL
L4467-02 - 1	VB-15242	9	0.0680	-	7470A	PPB	22 Oct 2020 10:52:38	U	SMPL
CCV76 - 1	CCV76	17561	5.0837	-	7470A	PPB	22 Oct 2020 10:54:54	U	SMPL
CCB76 - 1	CCB76	-100	0.0368	-	7470A	PPB	22 Oct 2020 10:57:09	U	SMPL
L4467-04 - 1	VNJ-240	-14	0.0614	-	7470A	PPB	22 Oct 2020 10:59:25	U	SMPL
L4467-06 - 1	357	-2	0.0648	-	7470A	PPB	22 Oct 2020 11:01:41	U	SMPL
L4468-02 - 1	72-11977	-10	0.0626	-	7470A	PPB	22 Oct 2020 11:03:56	U	SMPL
PB132534BL - 1	PBW	-5	0.0640	-	7470A	PPB	22 Oct 2020 11:05:53	U	SMPL
PB132534BS - 1	LCSW	13592	3.9495	-	7470A	PPB	22 Oct 2020 11:08:08	U	SMPL
L4469-01 - 1	FY-WATER	-64	0.0471	-	7470A	PPB	22 Oct 2020 11:10:24	U	SMPL
L4469-01DUP - 1	FY-WATERDUP	-17	0.0606	-	7470A	PPB	22 Oct 2020 11:12:41	U	SMPL
L4469-01MS - 1	FY-WATERMS	14439	4.1915	-	7470A	PPB	22 Oct 2020 11:14:58	U	SMPL
L4469-01MSD - 1	FY-WATERMSD	15277	4.4310	-	7470A	PPB	22 Oct 2020 11:17:14	U	SMPL
CCV77 - 1	CCV77	17606	5.0965	-	7470A	PPB	22 Oct 2020 11:19:30	U	SMPL
CCB77 - 1	CCB77	-131	0.0280	-	7470A	PPB	22 Oct 2020 11:21:46	U	SMPL
PB132533TB - 1	PB132507TB	-46	0.0523	-	7470A	PPB	22 Oct 2020 11:24:01	U	SMPL
PB132534TB - 1	PB132526TB	-9	0.0628	-	7470A	PPB	22 Oct 2020 11:26:16	U	SMPL
L4221-14LX5 - 1		-11	0.0623	-	7470A	PPB	22 Oct 2020 11:28:32	U	SMPL
L4221-14A - 1		13898	4.0369	-	7470A	PPB	22 Oct 2020 11:30:48	U	SMPL
L4469-01LX5 - 1		-96	0.0380	-	7470A	PPB	22 Oct 2020 11:33:04	U	SMPL
L4469-01A - 1		14676	4.2592	-	7470A	PPB	22 Oct 2020 11:35:19	U	SMPL
CCV78 - 1	CCV78	17658	5.1114	-	7470A	PPB	22 Oct 2020 11:37:35	U	SMPL
CCB78 - 1	CCB78	-126	0.0294	-	7470A	PPB	22 Oct 2020 11:39:52	U	SMPL