

ISM02.4


LB111650

INSTRUMENT ID: CV1

Linear

 μ Abs.:

38039



A= 0.0000e+000

B= 2.6487e-004

C= 3.1123e-002

Rho= 0.9997417

Accept=Accepted

slope
 $y = 1/m \cdot x + c$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.041	0.041	38	0.000	38					=
0.05	0.050					400					-7
0.20	0.200	0.187	-0.013	588	0.0 %	588					0
2.50	2.500	2.505	0.005	9339	0.0 %	9339					1
5.00	5.000	5.029	0.029	18868	0.0 %	18868					-2
7.50	7.500	7.332	-0.168	27565	0.0 %	27565					)
10.0	10.000	10.106	0.106	38039	0.0 %	38039					

LB111650

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	38	-		0 ISM02.4	PPB	10/22/2020 14:31	Std
	0.2 S0.2	588	-		0.2 ISM02.4	PPB	10/22/2020 14:34	Std
	2.5 S2.5	9339	-		2.5 ISM02.4	PPB	10/22/2020 14:36	Std
	5 S5.0	18868	-		5 ISM02.4	PPB	10/22/2020 14:38	Std
	7.5 S7.5	27565	-		7.5 ISM02.4	PPB	10/22/2020 14:43	Std
	10 S10.0	38039	-		10 ISM02.4	PPB	10/22/2020 14:48	Std
ICV	ICV	15713	4.193 -		ISM02.4	PPB	10/22/2020 14:51	SMPL
ICB	ICB	-27	0.024 -		ISM02.4	PPB	10/22/2020 14:53	SMPL
CCV079	CCV079	18469	4.923 -		ISM02.4	PPB	10/22/2020 14:55	SMPL
CCB079	CCB079	-17	0.0266 -		ISM02.4	PPB	10/22/2020 14:57	SMPL
PB132558BL	PBW558	9	0.0335 -		ISM02.4	PPB	10/22/2020 15:00	SMPL
L4320-10	MBEQ78	-26	0.0242 -		ISM02.4	PPB	10/22/2020 15:02	SMPL
PB132559BL	PBW559	13	0.0346 -		ISM02.4	PPB	10/22/2020 15:04	SMPL
L4416-01	MJLWH1	13	0.0346 -		ISM02.4	PPB	10/22/2020 15:06	SMPL
L4416-02	MJLWH5	-7	0.0293 -		ISM02.4	PPB	10/22/2020 15:09	SMPL
L4416-03	MJLWH7	13	0.0346 -		ISM02.4	PPB	10/22/2020 15:11	SMPL
L4416-04	MJLWJ2	-3	0.0303 -		ISM02.4	PPB	10/22/2020 15:13	SMPL
L4416-05	MJLWJ2D	-5	0.0298 -		ISM02.4	PPB	10/22/2020 15:15	SMPL
L4416-06	MJLWJ2S	4201	1.1438 -		ISM02.4	PPB	10/22/2020 15:20	SMPL
L4416-07	MJLWJ8	232	0.0926 -		ISM02.4	PPB	10/22/2020 15:23	SMPL
L4416-08	MJLWH8	890	0.2669 -		ISM02.4	PPB	10/22/2020 15:25	SMPL
L4416-09	MJLWH9	449	0.15 -		ISM02.4	PPB	10/22/2020 15:27	SMPL
L4416-10	MJLWJ5	115	0.0616 -		ISM02.4	PPB	10/22/2020 15:29	SMPL
L4416-11	MJLWJ6	-296	-0.0473 -		ISM02.4	PPB	10/22/2020 15:32	SMPL
L4416-12	MJLWK0	-129	-0.003 -		ISM02.4	PPB	10/22/2020 15:34	SMPL
L4416-13	MJLWK7	-46	0.0189 -		ISM02.4	PPB	10/22/2020 15:36	SMPL
PB132560BL	PBW560	-51	0.0176 -		ISM02.4	PPB	10/22/2020 15:38	SMPL
L4443-01	METYZ8	-57	0.016 -		ISM02.4	PPB	10/22/2020 15:41	SMPL
L4443-02	METZ13	-47	0.0187 -		ISM02.4	PPB	10/22/2020 15:43	SMPL
L4443-03	METZ18	-27	0.024 -		ISM02.4	PPB	10/22/2020 15:45	SMPL
CCV080	CCV080	17982	4.794 -		ISM02.4	PPB	10/22/2020 15:48	SMPL
CCB080	CCB080	-41	0.0203 -		ISM02.4	PPB	10/22/2020 15:50	SMPL
L4443-04	METZ32	98	0.0571 -		ISM02.4	PPB	10/22/2020 15:52	SMPL
L4443-05	METZ35	-34	0.0221 -		ISM02.4	PPB	10/22/2020 15:54	SMPL
L4443-06	METZ35D	-2	0.0306 -		ISM02.4	PPB	10/22/2020 15:57	SMPL
L4443-07	METZ35S	4220	1.1489 -		ISM02.4	PPB	10/22/2020 15:59	SMPL
L4443-08	METZ40	46	0.0433 -		ISM02.4	PPB	10/22/2020 16:01	SMPL
L4443-09	METZ44	59	0.0468 -		ISM02.4	PPB	10/22/2020 16:03	SMPL
L4443-10	METZ45	101	0.0579 -		ISM02.4	PPB	10/22/2020 16:06	SMPL
CCV081	CCV081	17739	4.7296 -		ISM02.4	PPB	10/22/2020 16:08	SMPL
CCB081	CCB081	-15	0.0272 -		ISM02.4	PPB	10/22/2020 16:10	SMPL