



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111652

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/22/2020	13:25
2	CAL2	1	Water	NA	NA	20.1	7.00	10/22/2020	13:26
3	CAL3	1	Water	NA	NA	20.1	10.03	10/22/2020	13:27
4	ICV	1	Water	NA	NA	20.2	7.00	10/22/2020	13:30
5	CCV1	1	Water	NA	NA	20.2	2.02	10/22/2020	13:32
6	L4494-01	1	Solid	20.03	20	22.6	9.28	10/22/2020	13:35
7	L4494-01DUP	1	Solid	20.04	20	22.7	9.29	10/22/2020	13:36
8	L4494-02	1	Solid	20.02	20	23.0	9.62	10/22/2020	13:37
9	CCV2	1	Water	NA	NA	20.3	12.03	10/22/2020	13:40

LB 111652

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph s14494 WorkList ID : 143574 Department : Wet-Chemistry Date : 10-22-2020 13:13:14

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/22/2020	Solid	L4494-01	~ pH	Cool 4 deg C	JPCL01	D12	AB-EH	10/22/2020	9045D
10/22/2020	Solid	L4494-02	~ pH	Cool 4 deg C	JPCL01	D12	AB-EN	10/22/2020	9045D

Date/Time 10/22/2020 13:20
 Received by: Jo (wet chem)
 Relinquished by: AML (wet chem)

Date/Time 10/22/2020 13:00
 Received by: AML (wet chem)
 Relinquished by: Jo (wet chem)