



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111657

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/22/2020	16:00
2	CAL2	1	Water	NA	NA	20.3	7.01	10/22/2020	16:01
3	CAL3	1	Water	NA	NA	20.3	10.02	10/22/2020	16:02
4	ICV	1	Water	NA	NA	20.4	6.99	10/22/2020	16:05
5	CCV1	1	Water	NA	NA	20.2	2.01	10/22/2020	16:06
6	L4494-03	1	Solid	20.03	20	22.6	9.28	10/22/2020	16:10
7	L4494-04	1	Solid	20.02	20	23.0	9.62	10/22/2020	16:11
8	L4496-04	1	Solid	20.03	20	23.1	7.59	10/22/2020	16:13
9	L4496-04DUP	1	Solid	20.04	20	23.2	7.60	10/22/2020	16:14
10	CCV2	1	Water	NA	NA	20.1	12.02	10/22/2020	16:15

VP 11657

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L4494 WorkList ID : 143573 Department : Wet-Chemistry Date : 10-22-2020 13:05:30

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/22/2020	Solid	L4494-03	Corrosivity	Cool 4 deg C	JPCL01	D12	AB-EH	10/22/2020	9045D
10/22/2020	Solid	L4494-04	Corrosivity	Cool 4 deg C	JPCL01	D12	AB-EN	10/22/2020	9045D
10/22/2020	Solid	L4496-04	Corrosivity	Cool 4 deg C	PSEG05	D22	SA-01-102220	10/22/2020	9045D

Date/Time 10.22.2020 13:40
 Received by: JP (Wet Chem)
 Relinquished by: KS (Wet Chem)

Date/Time 10.22.2020 14:30
 Received by: KS (Wet Chem)
 Relinquished by: JP