



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111658

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	21.1	4.01	10/22/2020	17:05
2	CAL2	1	Water	NA	NA	21.1	7.00	10/22/2020	17:06
3	CAL3	1	Water	NA	NA	21.1	10.03	10/22/2020	17:08
4	ICV	1	Water	NA	NA	21.2	7.00	10/22/2020	17:09
5	CCV1	1	Water	NA	NA	21.2	2.01	10/22/2020	17:10
6	L4486-07	1	Water	NA	NA	21.5	7.53	10/22/2020	17:11
7	L4486-07DUP	1	Water	NA	NA	21.6	7.54	10/22/2020	17:12
8	L4486-08	1	Water	NA	NA	21.6	6.29	10/22/2020	17:15
9	L4486-09	1	Water	NA	NA	21.7	7.34	10/22/2020	17:16
10	L4488-06	1	Water	NA	NA	22.2	10.03	10/22/2020	17:19
11	CCV2	1	Water	NA	NA	21.1	12.02	10/22/2020	17:20

28111658

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4486 WorkList ID : 143580 Department : Wet-Chemistry Date : 10-22-2020 16:48:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/22/2020	Water	L4486-07	F pH	Cool 4 deg C	PSEG01	D22	1836	10/22/2020	9040C
10/22/2020	Water	L4486-08	F pH	Cool 4 deg C	PSEG01	D22	1814	10/22/2020	9040C
10/22/2020	Water	L4486-09	F pH	Cool 4 deg C	PSEG01	D22	1806	10/22/2020	9040C
10/22/2020	Water	L4488-06	F pH	Cool 4 deg C	PSEG01	D31	1576	10/22/2020	9040C

Date/Time 10-22-2020 16:30
 Received by: JP (wetchem)
 Relinquished by: KS (wetchem)

Date/Time 10-22-2020 17:15
 Received by: KS (wetchem)
 Relinquished by: JP (wetchem) Chur