


LB111660

7470A

INSTRUMENT ID: CV2



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	-81	0.000	-81				
0.2	0.200	0.165	-0.035	459	0.0 %	459				
2.5	2.500	2.728	0.228	9308	0.0 %	9308				
5.0	5.000	4.940	-0.060	16947	0.0 %	16947				
7.5	7.500	7.132	-0.368	24514	0.0 %	24514				
10.0	10.000	10.225	0.225	35195	0.0 %	35195				

LB111660 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 23 Oct 2020 09:37:06				
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0.0 - 1	50	-81	-	0.0000	7470A	PPB	23 Oct 2020 09:52:15	S Std
0.2 - 1	50.2	459	-	0.2000	7470A	PPB	23 Oct 2020 09:54:32	S Std
2.5 - 1	52.5	9308	-	2.5000	7470A	PPB	23 Oct 2020 09:56:50	S Std
5.0 - 1	55	16947	-	5.0000	7470A	PPB	23 Oct 2020 09:59:06	S Std
7.5 - 1	57.5	24514	-	7.5000	7470A	PPB	23 Oct 2020 10:01:23	S Std
10.0 - 1	510	35195	-	10.0000	7470A	PPB	23 Oct 2020 10:03:40	S Std
ICV70 - 1	ICV70	14069	4.1069	-	7470A	PPB	23 Oct 2020 10:06:48	U SMPL
ICB70 - 1	ICB70	-38	0.0214	-	7470A	PPB	23 Oct 2020 10:11:52	U SMPL
CCV82 - 1	CCV82	17115	4.9891	-	7470A	PPB	23 Oct 2020 10:14:07	U SMPL
CCB82 - 1	CCB82	-131	-0.0055	-	7470A	PPB	23 Oct 2020 10:16:22	U SMPL
CRA - 1	CRA	444	0.1610	-	7470A	PPB	23 Oct 2020 10:18:37	U SMPL
HighStd - 1	HighStd	35715	10.3758	-	7470A	PPB	23 Oct 2020 10:20:52	U SMPL
ChkStd - 1	ChkStd	24379	7.0928	-	7470A	PPB	23 Oct 2020 10:23:07	U SMPL
PB132561BL - 1	PBW	-194	-0.0238	-	7470A	PPB	23 Oct 2020 10:25:24	U SMPL
PB132561BS - 1	LCSW	13964	4.0765	-	7470A	PPB	23 Oct 2020 10:27:41	U SMPL
L4225-14 - 1	SU-02-102120	16	0.0371	-	7470A	PPB	23 Oct 2020 10:29:57	U SMPL
L4225-14DUP - 1	SU-02-102120DUP	2	0.0330	-	7470A	PPB	23 Oct 2020 10:44:51	U SMPL
L4225-14MS - 1	SU-02-102120MS	14798	4.3181	-	7470A	PPB	23 Oct 2020 10:47:08	U SMPL
L4225-14MSD - 1	SU-02-102120MSD	15635	4.5605	-	7470A	PPB	23 Oct 2020 10:49:26	U SMPL
L4479-01 - 1	10-21A	-89	0.0067	-	7470A	PPB	23 Oct 2020 10:51:41	U SMPL
CCV83 - 1	CCV83	17242	5.0259	-	7470A	PPB	23 Oct 2020 10:53:57	U SMPL
CCB83 - 1	CCB83	-113	-0.0003	-	7470A	PPB	23 Oct 2020 10:56:12	U SMPL
L4479-02 - 1	10-21B	-11	0.0292	-	7470A	PPB	23 Oct 2020 10:58:28	U SMPL
L4479-03 - 1	10-21C	-2	0.0319	-	7470A	PPB	23 Oct 2020 11:00:43	U SMPL
L4479-04 - 1	10-21D	-11	0.0292	-	7470A	PPB	23 Oct 2020 11:02:58	U SMPL
L4479-05 - 1	10-21E	-7	0.0304	-	7470A	PPB	23 Oct 2020 11:05:14	U SMPL
L4479-06 - 1	10-21F	7	0.0345	-	7470A	PPB	23 Oct 2020 11:07:29	U SMPL
L4480-03 - 1	IS-SOIL-PILE	14	0.0365	-	7470A	PPB	23 Oct 2020 11:09:46	U SMPL
L4484-02 - 1	72-11998	-1	0.0321	-	7470A	PPB	23 Oct 2020 11:12:02	U SMPL
PB132561TB - 1	PB132536TB	7	0.0345	-	7470A	PPB	23 Oct 2020 11:14:19	U SMPL
L4225-14LX5 - 1		5	0.0339	-	7470A	PPB	23 Oct 2020 11:16:11	U SMPL
L4225-14A - 1		13975	4.0797	-	7470A	PPB	23 Oct 2020 11:18:26	U SMPL
CCV84 - 1	CCV84	17304	5.0438	-	7470A	PPB	23 Oct 2020 11:20:42	U SMPL
CCB84 - 1	CCB84	-116	-0.0012	-	7470A	PPB	23 Oct 2020 11:22:57	U SMPL