



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111675

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/23/2020	15:00
2	CAL2	1	Water	NA	NA	20.1	7.00	10/23/2020	15:01
3	CAL3	1	Water	NA	NA	20.1	10.03	10/23/2020	15:02
4	ICV	1	Water	NA	NA	20.1	6.98	10/23/2020	15:05
5	CCV1	1	Water	NA	NA	20.1	2.01	10/23/2020	15:06
6	L4500-02	1	Solid	20.02	20	22.8	8.71	10/23/2020	15:09
7	L4500-02DUP	1	Solid	20.03	20	22.9	8.72	10/23/2020	15:10
8	CCV2	1	Water	NA	NA	20.1	12.02	10/23/2020	15:11

783111675

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4500 WorkList ID : 143614 Department : Wet-Chemistry Date : 10-23-2020 14:43:28

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/23/2020	Solid	L4500-02	- pH	Cool 4 deg C	PSEG01	D31	11935	10/23/2020	9045D

Date/Time 10/23/2020 14:40
Received by: JH (Wet Chem)
Relinquished by: CR (SM)

Date/Time 10/23/2020 15:31
Received by: SE (SM)
Relinquished by: JH (Wet Chem)