

LB111676


7471B

INSTRUMENT ID: CV2

Linear

μ Abs.:

46917



A= 0.0000e+000

B= 2.1123e-004

C= -8.1337e-003

Rho= 0.9997584

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.002	0.002	49	0.000	49				
0.2	0.200	0.202	0.002	994	0.0 %	994				
2.5	2.500	2.477	-0.023	11764	0.0 %	11764				
5.0	5.000	4.952	-0.048	23484	0.0 %	23484				
7.5	7.500	7.665	0.165	36324	0.0 %	36324				
10.0	10.000	9.902	-0.098	46917	0.0 %	46917				

LB111676 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 23 Oct 2020 15:24:20

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	49	-	0.00007471B	PPB	23 Oct 2020 15:25:22	S	Std
0.2 - 1	50.2	994	-	0.20007471B	PPB	23 Oct 2020 15:30:09	S	Std
2.5 - 1	52.5	11764	-	2.50007471B	PPB	23 Oct 2020 15:32:26	S	Std
5.0 - 1	55	23484	-	5.00007471B	PPB	23 Oct 2020 15:34:43	S	Std
7.5 - 1	57.5	36324	-	7.50007471B	PPB	23 Oct 2020 15:37:00	S	Std
10.0 - 1	510	46917	-	10.00007471B	PPB	23 Oct 2020 15:44:28	S	Std
ICV72 - 1	ICV72	20504	4.3229	-7471B	PPB	23 Oct 2020 15:49:40	U	SMPL
ICB72 - 1	ICB72	-35	-0.0155	-7471B	PPB	23 Oct 2020 15:51:55	U	SMPL
CCV88 - 1	CCV88	23610	4.9790	-7471B	PPB	23 Oct 2020 15:54:10	U	SMPL
CCB88 - 1	CCB88	-40	-0.0166	-7471B	PPB	23 Oct 2020 15:56:25	U	SMPL
CRA - 1	CRA	969	0.1965	-7471B	PPB	23 Oct 2020 15:58:41	U	SMPL
HighStd - 1	HighStd	47099	9.9406	-7471B	PPB	23 Oct 2020 16:00:56	U	SMPL
ChkStd - 1	ChkStd	32684	6.8957	-7471B	PPB	23 Oct 2020 16:05:03	U	SMPL
PB132582BL - 1	PBS	-115	-0.0324	-7471B	PPB	23 Oct 2020 16:07:19	U	SMPL
PB132582BS - 1	LCSS	19491	4.1090	-7471B	PPB	23 Oct 2020 16:09:36	U	SMPL
L4225-10 - 1	SU-02-102120	2349	0.4880	-7471B	PPB	23 Oct 2020 16:11:53	U	SMPL
L4480-02 - 1	IS-SOIL-PILE	1966	0.4071	-7471B	PPB	23 Oct 2020 16:14:09	U	SMPL
L4480-02DUP - 1	IS-SOIL-PILEDUP	2261	0.4695	-7471B	PPB	23 Oct 2020 16:16:26	U	SMPL
L4480-02MS - 1	IS-SOIL-PILEMS	21926	4.6233	-7471B	PPB	23 Oct 2020 16:18:43	U	SMPL
L4480-02MSD - 1	IS-SOIL-PILEMSD	23404	4.9355	-7471B	PPB	23 Oct 2020 16:21:32	U	SMPL
CCV89 - 1	CCV89	23532	4.9625	-7471B	PPB	23 Oct 2020 16:23:47	U	SMPL
CCB89 - 1	CCB89	-37	-0.0159	-7471B	PPB	23 Oct 2020 16:26:02	U	SMPL
L4484-01 - 1	72-11998	2297	0.4771	-7471B	PPB	23 Oct 2020 16:28:17	U	SMPL
L4486-02 - 1	MS-SOIL	4234	0.8862	-7471B	PPB	23 Oct 2020 16:30:33	U	SMPL
L4488-05 - 1	202	59898	12.6441	-7471B	PPB	23 Oct 2020 16:32:48	U	SMPL
L4491-01 - 1	COMP-1	1567	0.3229	-7471B	PPB	23 Oct 2020 16:35:03	U	SMPL
L4491-02 - 1	COMP-2	1427	0.2933	-7471B	PPB	23 Oct 2020 16:37:21	U	SMPL
L4491-03 - 1	COMP-3	2230	0.4629	-7471B	PPB	23 Oct 2020 16:39:37	U	SMPL
L4496-01 - 1	SA-01-102220	937	0.1898	-7471B	PPB	23 Oct 2020 16:41:54	U	SMPL
L4480-02LX5 - 1		461	0.0892	-7471B	PPB	23 Oct 2020 16:44:11	U	SMPL
L4480-02A - 1		21588	4.5519	-7471B	PPB	23 Oct 2020 16:46:29	U	SMPL
L4488-05DLX2 - 1		30241	6.3797	-7471B	PPB	23 Oct 2020 16:48:44	U	SMPL
CCV90 - 1	CCV90	23560	4.9685	-7471B	PPB	23 Oct 2020 16:55:48	U	SMPL
CCB90 - 1	CCB90	-17	-0.0117	-7471B	PPB	23 Oct 2020 16:58:04	U	SMPL