



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB111679

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/23/2020	09:40
2	L4478-01		1	8	100.00	>212.0	>212.0	10/23/2020	10:10
3	L4478-01DUP		1	8	100.00	>212.0	>212.0	10/23/2020	10:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 111679

WorkList Name : fp102220 WorkList ID : 143557 Department : Wet-Chemistry Date : 10-22-2020 08:55:25

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/26/2020	Solid	L4478-01	B Flash Point	Cool 4 deg C	PSEG01	D21	HC-OIL	10/21/2020	1010A

Date/Time 10/23/20 09:15
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)

Date/Time 10/23/20 10:50
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)