



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB111680

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/23/2020	09:40
2	L4486-03		1	9	100.00	>212.0	>212.0	10/23/2020	11:25
3	L4486-03DUP		1	9	100.00	>212.0	>212.0	10/23/2020	11:50
4	L4486-05		1	8	100.00	>212.0	>212.0	10/23/2020	12:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 111680

WorkList Name : FP102220-1 WorkList ID : 143581 Department : Wet-Chemistry Date : 10-22-2020 17:08:44

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/29/2020	Water	L4486-03	Flash Point	Cool 4 deg C	PSEG01	D22	1839	10/22/2020	1010A
10/29/2020	Water	L4486-05	Flash Point	Cool 4 deg C	PSEG01	D22	1811	10/22/2020	1010A

Date/Time 10/23/20 10:55
 Received by: KS (Welcher)
 Relinquished by: KS (Welcher)

Date/Time 10/23/20 12:25
 Received by: KS (Welcher)
 Relinquished by: KS (Welcher)