



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB111681

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/26/2020	09:45
2	L4487-01		1	8	100.00	>212.0	>212.0	10/26/2020	10:20
3	L4487-02		1	8	100.00	>212.0	>212.0	10/26/2020	10:50
4	L4487-03		1	9	100.00	>212.0	>212.0	10/26/2020	11:15
5	L4487-04		1	9	100.00	>212.0	>212.0	10/26/2020	11:40
6	L4487-05		1	8	100.00	>212.0	>212.0	10/26/2020	12:10
7	L4487-06		1	9	100.00	>212.0	>212.0	10/26/2020	12:35
8	L4488-01		1	8	47.00	116.6	116.6	10/26/2020	13:05
9	L4488-01DUP		1	8	46.00	114.8	114.8	10/26/2020	13:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB111681

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP102320

WorkList ID : 143618

Department : Wet-Chemistry

Date : 10-23-2020 16:31:48

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/29/2020	Solid	L4487-01 ✓	Flash Point	Cool 4 deg C	PSEG01	D22	D-85	10/22/2020	1010A
10/29/2020	Solid	L4487-02 ✓	Flash Point	Cool 4 deg C	PSEG01	D22	D-88	10/22/2020	1010A
10/29/2020	Solid	L4487-03 ✓	Flash Point	Cool 4 deg C	PSEG01	D22	D-89	10/22/2020	1010A
10/29/2020	Solid	L4487-04 ✓	Flash Point	Cool 4 deg C	PSEG01	D22	D-90	10/22/2020	1010A
10/29/2020	Solid	L4487-05 ✓	Flash Point	Cool 4 deg C	PSEG01	D22	D-91	10/22/2020	1010A
10/29/2020	Solid	L4487-06 ✓	Flash Point	Cool 4 deg C	PSEG01	D22	D-92	10/22/2020	1010A
10/29/2020	Solid	L4488-01 ✓	Flash Point	Cool 4 deg C	PSEG01	D31	1618	10/22/2020	1010A

Date/Time 10/26/2020 09:15

Received by: KS (Wetchem)

Relinquished by: AH (Wetchem)

Date/Time

Received by:

Relinquished by:

10/26/2020 13:45
KS (Wetchem)
KS (Wetchem)