



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB111682

Slope : 97.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.000 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/23/2020	17:00
2	CAL2	1	Water	NA	NA	20.1	7.00	10/23/2020	17:01
3	CAL3	1	Water	NA	NA	20.1	10.03	10/23/2020	17:03
4	ICV	1	Water	NA	NA	20.2	7.00	10/23/2020	17:05
5	CCV1	1	Water	NA	NA	20.3	2.01	10/23/2020	17:09
6	L4219-24	1	Solid	20.03	20	23.0	9.27	10/23/2020	17:10
7	L4219-24DUP	1	Solid	20.04	20	23.1	9.28	10/23/2020	17:12
8	CCV2	1	Water	NA	NA	20.1	12.02	10/23/2020	17:15

LB 111682

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY L4219 WorkList ID : 143612 Department : Wet-Chemistry Date : 10-23-2020 12:37:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/23/2020	Solid	L4219-24	Corrosivity	Cool 4 deg C	PSEG05		CL-01-102320	10/23/2020	9045D

Date/Time 10/23/2020/6:10
 Received by: JB Wet Chem
 Relinquished by: CP SM

Date/Time 10/23/2020/8:00
 Received by: OP SM
 Relinquished by: JB Wet Chem