


LB 111685

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 3.1386e-004

C= 5.4807e-002

Rho= 0.9993806

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.048	0.048	-22	0.000	-22				
0.2	0.200	0.225	0.025	542	0.0 %	542				
2.5	2.500	2.276	-0.224	7078	0.0 %	7078				
5.0	5.000	5.211	0.211	16429	0.0 %	16429				
7.5	7.500	7.500	-0.000	23721	0.0 %	23721				
10.0	10.000	9.940	-0.060	31495	0.0 %	31495				

LB111685 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Oct 2020 08:31:11

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-22	-	0.0000	7470A	PPB	26 Oct 2020 08:41:00	S	Std
0.2 - 1	50.2	542	-	0.2000	7470A	PPB	26 Oct 2020 08:43:17	S	Std
2.5 - 1	52.5	7078	-	2.5000	7470A	PPB	26 Oct 2020 08:56:31	S	Std
5.0 - 1	55	16429	-	5.0000	7470A	PPB	26 Oct 2020 08:58:48	S	Std
7.5 - 1	57.5	23721	-	7.5000	7470A	PPB	26 Oct 2020 09:01:04	S	Std
10.0 - 1	510	31495	-	10.0000	7470A	PPB	26 Oct 2020 09:05:55	S	Std
ICV73 - 1	ICV73	13642	4.3365	-	7470A	PPB	26 Oct 2020 09:08:38	U	SMPL
ICB73 - 1	ICB73	-173	0.0005	-	7470A	PPB	26 Oct 2020 09:10:52	U	SMPL
CCV91 - 1	CCV91	16339	5.1830	-	7470A	PPB	26 Oct 2020 09:13:07	U	SMPL
CCB91 - 1	CCB91	-150	0.0077	-	7470A	PPB	26 Oct 2020 09:15:22	U	SMPL
CRA - 1	CRA	515	0.2164	-	7470A	PPB	26 Oct 2020 09:17:37	U	SMPL
HighStd - 1	HighStd	31710	10.0073	-	7470A	PPB	26 Oct 2020 09:19:52	U	SMPL
ChkStd - 1	ChkStd	22735	7.1904	-	7470A	PPB	26 Oct 2020 09:22:07	U	SMPL
PB132604BL - 1	PBW	-212	-0.0117	-	7470A	PPB	26 Oct 2020 09:24:24	U	SMPL
PB132604BS - 1	LCSW	13939	4.4297	-	7470A	PPB	26 Oct 2020 09:26:41	U	SMPL
L4494-03 - 1	AB-EH	-107	0.0212	-	7470A	PPB	26 Oct 2020 09:28:57	U	SMPL
L4494-04 - 1	AB-EN	-23	0.0476	-	7470A	PPB	26 Oct 2020 09:31:13	U	SMPL
L4496-04 - 1	SA-01-102220	-43	0.0413	-	7470A	PPB	26 Oct 2020 09:33:30	U	SMPL
L4496-04DUP - 1	SA-01-102220DUP	-15	0.0501	-	7470A	PPB	26 Oct 2020 09:35:47	U	SMPL
L4496-04MS - 1	SA-01-102220MS	13442	4.2737	-	7470A	PPB	26 Oct 2020 09:38:02	U	SMPL
CCV92 - 1	CCV92	16246	5.1538	-	7470A	PPB	26 Oct 2020 09:40:17	U	SMPL
CCB92 - 1	CCB92	-111	0.0200	-	7470A	PPB	26 Oct 2020 09:42:32	U	SMPL
L4496-04MSD - 1	SA-01-102220MSD	14073	4.4717	-	7470A	PPB	26 Oct 2020 09:44:49	U	SMPL
PB132605BL - 1	PBW	-108	0.0209	-	7470A	PPB	26 Oct 2020 09:47:04	U	SMPL
PB132605BS - 1	LCSW	13222	4.2047	-	7470A	PPB	26 Oct 2020 09:49:19	U	SMPL
L4434-10 - 1	16757-A	229	0.1267	-	7470A	PPB	26 Oct 2020 09:51:34	U	SMPL
L4486-03 - 1	1839	-44	0.0410	-	7470A	PPB	26 Oct 2020 09:53:51	U	SMPL
L4486-05 - 1	1811	-14	0.0504	-	7470A	PPB	26 Oct 2020 09:56:08	U	SMPL
L4486-07 - 1	1836	-4	0.0536	-	7470A	PPB	26 Oct 2020 09:58:24	U	SMPL
L4486-08 - 1	1814	-10	0.0517	-	7470A	PPB	26 Oct 2020 10:00:41	U	SMPL
L4486-09 - 1	1806	-14	0.0504	-	7470A	PPB	26 Oct 2020 10:02:58	U	SMPL
L4488-06 - 1	1576	6	0.0567	-	7470A	PPB	26 Oct 2020 10:05:13	U	SMPL
CCV93 - 1	CCV93	15885	5.0405	-	7470A	PPB	26 Oct 2020 10:07:29	U	SMPL
CCB93 - 1	CCB93	-124	0.0159	-	7470A	PPB	26 Oct 2020 10:09:44	U	SMPL
L4488-06DUP - 1	1576DUP	-22	0.0479	-	7470A	PPB	26 Oct 2020 10:12:00	U	SMPL
L4488-06MS - 1	1576MS	13658	4.3415	-	7470A	PPB	26 Oct 2020 10:14:16	U	SMPL
L4488-06MSD - 1	1576MSD	14497	4.6048	-	7470A	PPB	26 Oct 2020 10:16:33	U	SMPL
PB132604TB - 1	PB132562TB	-113	0.0193	-	7470A	PPB	26 Oct 2020 10:18:49	U	SMPL
PB132605TB - 1	PB132570TB	-27	0.0463	-	7470A	PPB	26 Oct 2020 10:21:05	U	SMPL
L4496-04LX5 - 1		-18	0.0492	-	7470A	PPB	26 Oct 2020 10:23:21	U	SMPL
L4496-04A - 1		13695	4.3531	-	7470A	PPB	26 Oct 2020 10:25:38	U	SMPL
L4488-06LX5 - 1		-99	0.0237	-	7470A	PPB	26 Oct 2020 10:27:54	U	SMPL
L4488-06A - 1		13616	4.3283	-	7470A	PPB	26 Oct 2020 10:30:11	U	SMPL
CCV94 - 1	CCV94	16030	5.0860	-	7470A	PPB	26 Oct 2020 10:32:27	U	SMPL
CCB94 - 1	CCB94	-155	0.0062	-	7470A	PPB	26 Oct 2020 10:34:42	U	SMPL