



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111691

Slope : 97.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/26/2020	14:45
2	CAL2	1	Water	NA	NA	20.1	7.00	10/26/2020	14:46
3	CAL3	1	Water	NA	NA	20.1	10.02	10/26/2020	14:47
4	ICV	1	Water	NA	NA	20.1	6.98	10/26/2020	14:48
5	CCV1	1	Water	NA	NA	20.2	2.01	10/26/2020	14:49
6	L4516-01	1	Water	NA	NA	16.8	5.84	10/26/2020	14:50
7	L4516-04	1	Water	NA	NA	17.0	5.81	10/26/2020	14:52
8	L4520-01	1	Water	NA	NA	16.8	8.42	10/26/2020	14:53
9	L4520-01DUP	1	Water	NA	NA	16.9	8.43	10/26/2020	14:54
10	CCV2	1	Water	NA	NA	20.1	12.02	10/26/2020	15:00

28111691

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4216 WorkList ID : 143644 Department : Wet-Chemistry Date : 10-26-2020 14:36:52

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/26/2020	Water	L4516-01-11	pH	Cool 4 deg C	PSEG01	D41	FRAC-TANK-1	10/26/2020	9040C
10/26/2020	Water	L4516-04	pH	Cool 4 deg C	PSEG01	D41	FRAC-TANK-2	10/26/2020	9040C
10/26/2020	Water	L4520-01	pH	Cool 4 deg C	PSEG01	D41	PS-FRAC	10/26/2020	9040C

Date/Time 10/26/2020 14:40
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)

Date/Time 10/26/2020 16:00
 Received by: KS (Wetchem)
 Relinquished by: KS (Wetchem)