


LB111692

7471B

INSTRUMENT ID: CV1

 μ Abs.:

52799



Linear

A= 0.0000e+000

B= 1.8886e-004

C= 4.5409e-002

Rho= 0.9999004

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.046	0.046	3	0.000	3				
0.2	0.200	0.231	0.031	982	0.0 %	982				
2.5	2.500	2.439	-0.061	12673	0.0 %	12673				
5.0	5.000	4.918	-0.082	25799	0.0 %	25799				
7.5	7.500	7.549	0.049	39732	0.0 %	39732				
10.0	10.000	10.017	0.017	52799	0.0 %	52799				

LB111692 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 26 Oct 2020 15:02:33

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	3	-	0.0000	7471B	PPB	26 Oct 2020 15:22:39	S	Std
0.2 - 1	50.2	982	-	0.2000	7471B	PPB	26 Oct 2020 15:24:56	S	Std
2.5 - 1	52.5	12673	-	2.5000	7471B	PPB	26 Oct 2020 15:27:13	S	Std
5.0 - 1	55	25799	-	5.0000	7471B	PPB	26 Oct 2020 15:29:29	S	Std
7.5 - 1	57.5	39732	-	7.5000	7471B	PPB	26 Oct 2020 15:31:46	S	Std
10.0 - 1	510	52799	-	10.0000	7471B	PPB	26 Oct 2020 15:34:01	S	Std
ICV74 - 1	ICV74	21153	4.0404	-7471B	PPB	PPB	26 Oct 2020 15:37:56	U	SMPL
ICB74 - 1	ICB74	46	0.0541	-7471B	PPB	PPB	26 Oct 2020 15:42:47	U	SMPL
CCV97 - 1	CCV97	25986	4.9532	-7471B	PPB	PPB	26 Oct 2020 15:45:02	U	SMPL
CCB97 - 1	CCB97	-52	0.0356	-7471B	PPB	PPB	26 Oct 2020 15:47:16	U	SMPL
CRA - 1	CRA	969	0.2284	-7471B	PPB	PPB	26 Oct 2020 15:49:32	U	SMPL
HighStd - 1	HighStd	53398	10.1303	-7471B	PPB	PPB	26 Oct 2020 15:51:47	U	SMPL
ChkStd - 1	ChkStd	38635	7.3421	-7471B	PPB	PPB	26 Oct 2020 15:54:02	U	SMPL
PB132632BL - 1	PBS	-134	0.0201	-7471B	PPB	PPB	26 Oct 2020 15:56:19	U	SMPL
PB132632BS - 1	LCSS	19786	3.7822	-7471B	PPB	PPB	26 Oct 2020 15:58:36	U	SMPL
L4219-17 - 1	CL-01-102320	1193	0.2707	-7471B	PPB	PPB	26 Oct 2020 16:00:52	U	SMPL
L4500-02 - 1	11935	4493	0.8940	-7471B	PPB	PPB	26 Oct 2020 16:03:09	U	SMPL
L4501-01 - 1	COMP-1	2535	0.5242	-7471B	PPB	PPB	26 Oct 2020 16:05:25	U	SMPL
L4501-02 - 1	COMP-2	1485	0.3259	-7471B	PPB	PPB	26 Oct 2020 16:07:42	U	SMPL
L4501-03 - 1	COMP-3	1818	0.3888	-7471B	PPB	PPB	26 Oct 2020 16:09:57	U	SMPL
CCV98 - 1	CCV98	25830	4.9237	-7471B	PPB	PPB	26 Oct 2020 16:12:13	U	SMPL
CCB98 - 1	CCB98	-86	0.0292	-7471B	PPB	PPB	26 Oct 2020 16:14:28	U	SMPL
L4508-01 - 1	PM-WC-C	820	0.2003	-7471B	PPB	PPB	26 Oct 2020 16:16:43	U	SMPL
L4510-16 - 1	COMP-1	2376	0.4941	-7471B	PPB	PPB	26 Oct 2020 16:18:58	U	SMPL
L4510-16DUP - 1	COMP-1DUP	2168	0.4549	-7471B	PPB	PPB	26 Oct 2020 16:21:13	U	SMPL
L4510-16MS - 1	COMP-1MS	21298	4.0678	-7471B	PPB	PPB	26 Oct 2020 16:23:29	U	SMPL
L4510-16MSD - 1	COMP-1MSD	21081	4.0268	-7471B	PPB	PPB	26 Oct 2020 16:25:45	U	SMPL
L4510-17 - 1	COMP-2	1276	0.2864	-7471B	PPB	PPB	26 Oct 2020 16:28:01	U	SMPL
L4510-18 - 1	COMP-3	3174	0.6449	-7471B	PPB	PPB	26 Oct 2020 16:30:17	U	SMPL
L4510-19 - 1	COMP-4	4548	0.9044	-7471B	PPB	PPB	26 Oct 2020 16:32:34	U	SMPL
L4510-20 - 1	COMP-5	4887	0.9684	-7471B	PPB	PPB	26 Oct 2020 16:34:51	U	SMPL
L4510-21 - 1	COMP-6	4310	0.8594	-7471B	PPB	PPB	26 Oct 2020 16:37:06	U	SMPL
CCV99 - 1	CCV99	26204	4.9944	-7471B	PPB	PPB	26 Oct 2020 16:39:21	U	SMPL
CCB99 - 1	CCB99	-88	0.0288	-7471B	PPB	PPB	26 Oct 2020 16:41:37	U	SMPL
L4510-16LX5 - 1		423	0.1253	-7471B	PPB	PPB	26 Oct 2020 16:43:52	U	SMPL
L4510-16A - 1		24911	4.7502	-7471B	PPB	PPB	26 Oct 2020 16:46:08	U	SMPL
CCV01 - 1	CCV01	25482	4.8580	-7471B	PPB	PPB	26 Oct 2020 16:48:23	U	SMPL
CCB01 - 1	CCB01	-108	0.0250	-7471B	PPB	PPB	26 Oct 2020 16:50:39	U	SMPL