



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB111693

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/27/2020	09:32
2	L4516-02		1	8	100.00	>212.0	>212.0	10/27/2020	10:05
3	L4516-02DUP		1	8	100.00	>212.0	>212.0	10/27/2020	10:35
4	L4516-05		1	9	100.00	>212.0	>212.0	10/27/2020	11:10
5	L4520-02		1	9	100.00	>212.0	>212.0	10/27/2020	11:40

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB111693

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP102620

WorkList ID : 143645

Department : Wet-Chemistry

Date : 10-26-2020 16:16:45

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/29/2020	Water	L4516-02	Flash Point	Cool 4 deg C	PSEG01	D41	FRAC-TANK-1	10/26/2020	1010A
10/29/2020	Water	L4516-05	Flash Point	Cool 4 deg C	PSEG01	D41	FRAC-TANK-2	10/26/2020	1010A
10/29/2020	Water	L4520-02	Flash Point	Cool 4 deg C	PSEG01	D41	PS-FRAC	10/26/2020	1010A

Date/Time 10/27/20 09:00
Received by: KS (wekchen)
Relinquished by: KS (wekchen)

Date/Time 10/27/20 11:45
Received by: KS (wekchen)
Relinquished by: KS (wekchen)