


LB111699

ISM02.4

INSTRUMENT ID: CV2

 $\mu$  Abs.:

35580



Linear

A= 0.0000e+000

B= 2.8017e-004

C= 3.0915e-002

Rho= 0.9999849

Accept=Accepted

*slope*  
*x-intercept*

| Std ID | Conc.  | Calc. | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | 2/0D |
|--------|--------|-------|--------|-------|------------|-------|-------|-------|-------|-------|------|
| 0.00   | 0.000  | 0.035 | 0.035  | 15    | 0.000      | 15    |       |       |       |       | 1    |
| 0.05   | 0.050  |       |        |       |            | 400   |       |       |       |       | 1    |
| 0.20   | 0.200  | 0.181 | -0.019 | 537   | 0.0 %      | 537   |       |       |       |       | 10   |
| 2.50   | 2.500  | 2.483 | -0.017 | 8753  | 0.0 %      | 8753  |       |       |       |       | 1    |
| 5.00   | 5.000  | 4.983 | -0.017 | 17675 | 0.0 %      | 17675 |       |       |       |       | 0    |
| 7.50   | 7.500  | 7.518 | 0.018  | 26723 | 0.0 %      | 26723 |       |       |       |       | 0    |
| 10.0   | 10.000 | 9.999 | -0.001 | 35580 | 0.0 %      | 35580 |       |       |       |       | 0    |

LB111699  
INSTRUMENT ID : CV1

| Sample ID  | Extended ID | $\mu$ Abs. | Conc.    | Std Conc | Method      | Units | Date            | Type |
|------------|-------------|------------|----------|----------|-------------|-------|-----------------|------|
|            | 0 S0.0      | 15         | -        |          | 0 ISM02.4   | PPB   | 10/27/2020 8:13 | Std  |
|            | 0.2 S0.2    | 537        | -        |          | 0.2 ISM02.4 | PPB   | 10/27/2020 8:16 | Std  |
|            | 2.5 S2.5    | 8753       | -        |          | 2.5 ISM02.4 | PPB   | 10/27/2020 8:18 | Std  |
|            | 5 S5.0      | 17675      | -        |          | 5 ISM02.4   | PPB   | 10/27/2020 8:22 | Std  |
|            | 7.5 S7.5    | 26723      | -        |          | 7.5 ISM02.4 | PPB   | 10/27/2020 8:25 | Std  |
|            | 10 S10.0    | 35580      | -        |          | 10 ISM02.4  | PPB   | 10/27/2020 8:29 | Std  |
| ICV        | ICV         | 14688      | 4.1461 - |          | ISM02.4     | PPB   | 10/27/2020 8:32 | SMPL |
| ICB        | ICB         | -78        | 0.0091 - |          | ISM02.4     | PPB   | 10/27/2020 8:34 | SMPL |
| CCV002     | CCV002      | 17769      | 5.0093 - |          | ISM02.4     | PPB   | 10/27/2020 8:37 | SMPL |
| CCB002     | CCB002      | -81        | 0.0082 - |          | ISM02.4     | PPB   | 10/27/2020 8:39 | SMPL |
| PB132631BL | PBW         | 78         | 0.0528 - |          | ISM02.4     | PPB   | 10/27/2020 8:44 | SMPL |
| PB132631TB | LEB631      | 211        | 0.09 -   |          | ISM02.4     | PPB   | 10/27/2020 8:46 | SMPL |
| L4509-01   | MC0AQ2      | 139        | 0.0699 - |          | ISM02.4     | PPB   | 10/27/2020 8:48 | SMPL |
| L4509-02   | MC0AQ5      | 144        | 0.0713 - |          | ISM02.4     | PPB   | 10/27/2020 8:50 | SMPL |
| L4509-03   | MC0AQ6      | 155        | 0.0743 - |          | ISM02.4     | PPB   | 10/27/2020 8:53 | SMPL |
| L4509-04   | MC0AQ6D     | 126        | 0.0662 - |          | ISM02.4     | PPB   | 10/27/2020 8:55 | SMPL |
| L4509-05   | MC0AQ6S     | 4045       | 1.1642 - |          | ISM02.4     | PPB   | 10/27/2020 9:00 | SMPL |
| CCV003     | CCV003      | 17861      | 5.035 -  |          | ISM02.4     | PPB   | 10/27/2020 9:02 | SMPL |
| CCB003     | CCB003      | -26        | 0.0236 - |          | ISM02.4     | PPB   | 10/27/2020 9:05 | SMPL |