


LB111700

7470A

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.6331e-004

C= -2.2244e-002

Rho= 0.9999369

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.024	0.024	174	0.000	174				
0.2	0.200	0.216	0.016	905	0.0 %	905				
2.5	2.500	2.482	-0.018	9512	0.0 %	9512				
5.0	5.000	4.922	-0.078	18777	0.0 %	18777				
7.5	7.500	7.556	0.056	28779	0.0 %	28779				
10.0	10.000	10.000	0.000	38063	0.0 %	38063				

LB111700 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 27 Oct 2020 10:11:33

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	174	-	0.00007470A	PPB	27 Oct 2020 10:37:17	S	Std
0.2 - 1	50.2	905	-	0.20007470A	PPB	27 Oct 2020 10:39:34	S	Std
2.5 - 1	52.5	9512	-	2.50007470A	PPB	27 Oct 2020 10:41:51	S	Std
5.0 - 1	55	18777	-	5.00007470A	PPB	27 Oct 2020 10:44:08	S	Std
7.5 - 1	57.5	28779	-	7.50007470A	PPB	27 Oct 2020 10:46:25	S	Std
10.0 - 1	510	38063	-	10.00007470A	PPB	27 Oct 2020 10:51:12	S	Std
ICV75 - 1	ICV75	14054	3.6784	-7470A	PPB	27 Oct 2020 10:53:52	U	SMPL
ICB75 - 1	ICB75	34	-0.0133	-7470A	PPB	27 Oct 2020 10:56:07	U	SMPL
CCV04 - 1	CCV04	18875	4.9478	-7470A	PPB	27 Oct 2020 10:58:22	U	SMPL
CCB04 - 1	CCB04	51	-0.0088	-7470A	PPB	27 Oct 2020 11:00:37	U	SMPL
CRA - 1	CRA	840	0.1989	-7470A	PPB	27 Oct 2020 11:02:52	U	SMPL
HighStd - 1	HighStd	37866	9.9484	-7470A	PPB	27 Oct 2020 11:05:07	U	SMPL
ChkStd - 1	ChkStd	27134	7.1225	-7470A	PPB	27 Oct 2020 11:07:22	U	SMPL
PB132630BL - 1	PBW	-134	-0.0575	-7470A	PPB	27 Oct 2020 11:09:39	U	SMPL
PB132630BS - 1	LCSW	14115	3.6944	-7470A	PPB	27 Oct 2020 11:14:31	U	SMPL
L4219-24 - 1	CL-01-102320	2	-0.0217	-7470A	PPB	27 Oct 2020 11:16:47	U	SMPL
L4219-24DUP - 1	CL-01-102320DUP	86	0.0004	-7470A	PPB	27 Oct 2020 11:19:03	U	SMPL
L4219-24MS - 1	CL-01-102320MS	14265	3.7339	-7470A	PPB	27 Oct 2020 11:24:04	U	SMPL
L4219-24MSD - 1	CL-01-102320MSD	14383	3.7650	-7470A	PPB	27 Oct 2020 11:26:21	U	SMPL
L4434-08 - 1	40978-S	6	-0.0207	-7470A	PPB	27 Oct 2020 11:28:36	U	SMPL
CCV05 - 1	CCV05	18497	4.8483	-7470A	PPB	27 Oct 2020 11:30:51	U	SMPL
CCB05 - 1	CCB05	44	-0.0107	-7470A	PPB	27 Oct 2020 11:33:06	U	SMPL
L4434-09 - 1	16757-S	743	0.1734	-7470A	PPB	27 Oct 2020 11:35:22	U	SMPL
L4500-03 - 1	11935	137	0.0138	-7470A	PPB	27 Oct 2020 11:37:37	U	SMPL
L4510-22 - 1	COMP-1	478	0.1036	-7470A	PPB	27 Oct 2020 11:39:52	U	SMPL
L4510-23 - 1	COMP-2	417	0.0876	-7470A	PPB	27 Oct 2020 11:42:08	U	SMPL
L4510-24 - 1	COMP-3	99	0.0038	-7470A	PPB	27 Oct 2020 11:44:25	U	SMPL
L4510-25 - 1	COMP-4	87	0.0007	-7470A	PPB	27 Oct 2020 11:46:41	U	SMPL
L4510-26 - 1	COMP-5	89	0.0012	-7470A	PPB	27 Oct 2020 11:48:58	U	SMPL
L4510-27 - 1	COMP-6	126	0.0109	-7470A	PPB	27 Oct 2020 11:51:15	U	SMPL
PB132633BL - 1	PBW	77	-0.0020	-7470A	PPB	27 Oct 2020 11:53:31	U	SMPL
PB132633BS - 1	LCSW	13882	3.6331	-7470A	PPB	27 Oct 2020 11:55:47	U	SMPL
CCV06 - 1	CCV06	18467	4.8404	-7470A	PPB	27 Oct 2020 11:58:02	U	SMPL
CCB06 - 1	CCB06	32	-0.0138	-7470A	PPB	27 Oct 2020 12:00:18	U	SMPL
L4511-01 - 1	OB-7	55	-0.0078	-7470A	PPB	27 Oct 2020 12:02:33	U	SMPL
L4511-02 - 1	OB-9	216	0.0346	-7470A	PPB	27 Oct 2020 12:04:49	U	SMPL
L4511-02DUP - 1	OB-9DUP	211	0.0333	-7470A	PPB	27 Oct 2020 12:07:04	U	SMPL
L4511-02MS - 1	OB-9MS	14663	3.8387	-7470A	PPB	27 Oct 2020 12:09:20	U	SMPL
L4511-02MSD - 1	OB-9MSD	13958	3.6531	-7470A	PPB	27 Oct 2020 12:11:36	U	SMPL
L4516-01 - 1	FRAC-TANK-1	-58	-0.0375	-7470A	PPB	27 Oct 2020 12:13:52	U	SMPL
L4516-04 - 1	FRAC-TANK-2	0	-0.0222	-7470A	PPB	27 Oct 2020 12:16:08	U	SMPL
L4520-01 - 1	PS-FRAC	22	-0.0165	-7470A	PPB	27 Oct 2020 12:18:25	U	SMPL
CCV07 - 1	CCV07	18685	4.8978	-7470A	PPB	27 Oct 2020 12:20:42	U	SMPL
CCB07 - 1	CCB07	69	-0.0041	-7470A	PPB	27 Oct 2020 12:22:57	U	SMPL
PB132630TB - 1	PB132598TB	104	0.0051	-7470A	PPB	27 Oct 2020 12:25:12	U	SMPL
L4219-24LX5 - 1		36	-0.0128	-7470A	PPB	27 Oct 2020 12:27:28	U	SMPL
L4219-24A - 1		14004	3.6652	-7470A	PPB	27 Oct 2020 12:29:44	U	SMPL
L4511-02 - 1		-47	-0.0346	-7470A	PPB	27 Oct 2020 12:32:00	U	SMPL
L4511-02 - 1		14591	3.8198	-7470A	PPB	27 Oct 2020 12:34:15	U	SMPL
CCV08 - 1	CCV08	18852	4.9418	-7470A	PPB	27 Oct 2020 12:36:31	U	SMPL
CCB08 - 1	CCB08	-63	-0.0388	-7470A	PPB	27 Oct 2020 12:38:46	U	SMPL