



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 10/28/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 16:00
In Date: 10/27/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 10/28/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB111707

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
L4507-01	COBL2	1	1.14	9.77	8.81	88.9	
L4507-02	COBL3	2	1.18	9.53	8.31	85.4	
L4507-03	COBM1	3	1.16	9.98	7.6	73.0	
L4507-04	COBM2	4	1.12	9.97	8.1	78.9	
L4507-05	COBL8	5	1.12	9.67	8.09	81.5	
L4507-06	COBL9	6	1.13	9.83	8.58	85.6	
L4507-07	COBM0	7	1.13	9.57	8.53	87.7	
L4507-08	COBM3	8	1.15	9.85	8.69	86.7	
L4507-09	COBM4	9	1.14	9.96	8.57	84.2	
L4507-10	COBM5	10	1.15	9.68	7.76	77.5	
L4507-11	COBM6	11	1.14	9.67	7.82	78.3	
L4507-12	COBM7	12	1.13	9.84	8.19	81.1	
L4507-13	COBN0	13	1.14	9.63	7.82	78.7	
L4507-14	COBN1	14	1.15	9.65	7.81	78.4	
L4507-15	COBN2	15	1.14	9.83	9.34	94.4	
L4507-16	COBN3	16	1.15	9.75	8.08	80.6	
L4507-17	COBN4	17	1.16	9.7	8.4	84.8	
L4507-18	COBN5	18	1.13	9.92	8.12	79.5	
L4507-19	COBN6	19	1.13	9.54	7.67	77.8	
L4507-20	COBN8	20	1.16	9.65	8.13	82.1	
L4507-21	COBN8MS	21	1.16	9.65	8.13	82.1	
L4507-22	COBN8MSD	22	1.16	9.65	8.13	82.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$