

LB111708

INSTRUMENT ID: CV1

ISM02-4


AB SFAM01.1

10/27/18

Linear

 μ Abs.:

38796



A= 0.0000e+000

B= 2.5746e-004

C= 2.2746e-002

Rho= 0.9999388

Accept=Accepted

slope
y = intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%b
0.00	0.000	0.036	0.036	50	0.000	50					=
0.05	0.050					400					16
0.20	0.200	0.232	0.032	811	0.0 %	811					-3
2.50	2.500	2.422	-0.078	9317	0.0 %	9317					-1
5.00	5.000	4.973	-0.027	19228	0.0 %	19228					0
7.50	7.500	7.527	0.027	29147	0.0 %	29147					0
10.0	10.000	10.011	0.011	38796	0.0 %	38796					0

LB111708

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	50	-		0 SFAM01.0	PPB	10/27/2020 15:10	Std
	0.2 S0.2	811	-		0.2 SFAM01.0	PPB	10/27/2020 15:12	Std
	2.5 S2.5	9317	-		2.5 SFAM01.0	PPB	10/27/2020 15:14	Std
	5 S5.0	19228	-		5 SFAM01.0	PPB	10/27/2020 15:17	Std
	7.5 S7.5	29147	-		7.5 SFAM01.0	PPB	10/27/2020 15:19	Std
	10 S10.0	38796	-		10 SFAM01.0	PPB	10/27/2020 15:21	Std
ICV	ICV	14939	3.8689 -		SFAM01.0	PPB	10/27/2020 15:27	SMPL
ICB	ICB	-132	-0.0112 -		SFAM01.0	PPB	10/27/2020 15:29	SMPL
CCV011	CCV011	19704	5.0957 -		SFAM01.0	PPB	10/27/2020 15:31	SMPL
CCB011	CCB011	-113	-0.0063 -		SFAM01.0	PPB	10/27/2020 15:34	SMPL
PBS001	PBS001	-20	0.0176 -		SFAM01.0	PPB	10/27/2020 15:36	SMPL
PBS002	PBS002	-21	0.0173 -		SFAM01.0	PPB	10/27/2020 15:38	SMPL
L4485-01	MDL-SOIL-01	111	0.0513 -		SFAM01.0	PPB	10/27/2020 15:48	SMPL
L4485-02	MDL-SOIL-02	98	0.048 -		SFAM01.0	PPB	10/27/2020 15:50	SMPL
CCV012	CCV012	19267	4.9832 -		SFAM01.0	PPB	10/27/2020 15:52	SMPL
CCB012	CCB012	-108	-0.0051 -		SFAM01.0	PPB	10/27/2020 15:55	SMPL