

LB111715


7470A

INSTRUMENT ID: CY1

Linear

 μ Abs.:

35208



A= 0.0000e+000

B= 2.8516e-004

C= 1.1489e-002

Rho= 0.9999461

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.006	0.006	-21	0.000	-21				
0.2	0.200	0.225	0.025	747	0.0 %	747				
2.5	2.500	2.478	-0.022	8650	0.0 %	8650				
5.0	5.000	5.011	0.011	17531	0.0 %	17531				
7.5	7.500	7.430	-0.070	26014	0.0 %	26014				
10.0	10.000	10.051	0.051	35208	0.0 %	35208				

LB111715 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 28 Oct 2020 10:00:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-21	-	0.00007470A	PPB		28 Oct 2020 10:07:47	S	Std
0.2 - 1	80.2	747	-	0.20007470A	PPB		28 Oct 2020 10:10:04	S	Std
2.5 - 1	52.5	8650	-	2.50007470A	PPB		28 Oct 2020 10:12:21	S	Std
5.0 - 1	55	17531	-	5.00007470A	PPB		28 Oct 2020 10:14:38	S	Std
7.5 - 1	57.5	26014	-	7.50007470A	PPB		28 Oct 2020 10:16:54	S	Std
10.0 - 1	510	35208	-	10.00007470A	PPB		28 Oct 2020 10:21:40	S	Std
ICV76 - 1	ICV76	12893	3.6881	-7470A	PPB		28 Oct 2020 10:24:14	U	SMPL
ICB76 - 1	ICB76	-160	-0.0341	-7470A	PPB		28 Oct 2020 10:26:29	U	SMPL
CCV13 - 1	CCV13	18012	5.1478	-7470A	PPB		28 Oct 2020 10:28:44	U	SMPL
CCB13 - 1	CCB13	-130	-0.0256	-7470A	PPB		28 Oct 2020 10:30:59	U	SMPL
CRA - 1	CRA	722	0.2174	-7470A	PPB		28 Oct 2020 10:33:14	U	SMPL
HighStd - 1	HighStd	35632	10.1724	-7470A	PPB		28 Oct 2020 10:35:29	U	SMPL
ChkStd - 1	ChkStd	25773	7.3610	-7470A	PPB		28 Oct 2020 10:37:44	U	SMPL
PB132662BL - 1	PBW	-182	-0.0404	-7470A	PPB		28 Oct 2020 10:40:01	U	SMPL
PB132662BS - 1	LCSW	13429	3.8409	-7470A	PPB		28 Oct 2020 10:42:18	U	SMPL
L4516-02 - 1	FRAC-TANK-1	-112	-0.0204	-7470A	PPB		28 Oct 2020 10:44:34	U	SMPL
L4516-05 - 1	FRAC-TANK-2	-29	0.0032	-7470A	PPB		28 Oct 2020 10:46:50	U	SMPL
L4520-02 - 1	PS-FRAC	-20	0.0058	-7470A	PPB		28 Oct 2020 10:49:07	U	SMPL
L4520-02DUP - 1	PS-FRACDUP	-1	0.0112	-7470A	PPB		28 Oct 2020 10:51:24	U	SMPL
L4520-02MS - 1	PS-FRACMS	13627	3.8974	-7470A	PPB		28 Oct 2020 10:53:39	U	SMPL
CCV14 - 1	CCV14	18001	5.1447	-7470A	PPB		28 Oct 2020 10:55:54	U	SMPL
CCB14 - 1	CCB14	-137	-0.0276	-7470A	PPB		28 Oct 2020 10:58:09	U	SMPL
L4520-02MSD - 1	PS-FRACMSD	15705	4.4900	-7470A	PPB		28 Oct 2020 11:00:24	U	SMPL
PB132662TB - 1	PB132639TB	-100	-0.0170	-7470A	PPB		28 Oct 2020 11:02:40	U	SMPL
L4520-02LX5 - 1		-43	-0.0008	-7470A	PPB		28 Oct 2020 11:04:55	U	SMPL
L4520-02A - 1		14174	4.0534	-7470A	PPB		28 Oct 2020 11:07:11	U	SMPL
CCV15 - 1	CCV15	17961	5.1333	-7470A	PPB		28 Oct 2020 11:09:26	U	SMPL
CCB15 - 1	CCB15	-122	-0.0233	-7470A	PPB		28 Oct 2020 11:11:42	U	SMPL