


LB111722

SFAM01.0

INSTRUMENT ID: CV1

 $\mu$  Abs.:

34745



Linear

A= 0.0000e+000

B= 2.8646e-004

C= -4.5626e-003

Rho= 0.9999223

Accept=Accepted

*Slope*  
*y-intercept*

| Std ID | Conc.  | Calc.  | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 | %D |
|--------|--------|--------|--------|-------|------------|-------|-------|-------|-------|-------|----|
| 0.00   | 0.000  | -0.002 | -0.002 | 10    | 0.000      | 10    |       |       |       |       | =  |
| 0.05   | 0.050  |        |        |       |            | 400   |       |       |       |       | -4 |
| 0.20   | 0.200  | 0.208  | 0.008  | 743   | 0.0 %      | 743   |       |       |       |       | -3 |
| 2.50   | 2.500  | 2.437  | -0.063 | 8522  | 0.0 %      | 8522  |       |       |       |       | 1  |
| 5.00   | 5.000  | 5.061  | 0.061  | 17684 | 0.0 %      | 17684 |       |       |       |       | 1  |
| 7.50   | 7.500  | 7.547  | 0.047  | 26362 | 0.0 %      | 26362 |       |       |       |       | -1 |
| 10.0   | 10.000 | 9.948  | -0.052 | 34745 | 0.0 %      | 34745 |       |       |       |       |    |

LB111722

INSTRUMENT ID : CV1

| Sample ID | Extended ID  | $\mu$ Abs. | Conc.     | Std Conc | Method       | Units | Date             | Type |
|-----------|--------------|------------|-----------|----------|--------------|-------|------------------|------|
|           | 0 S0.0       | 10         | -         |          | 0 SFAM01.0   | PPB   | 10/28/2020 12:12 | Std  |
|           | 0.2 S0.2     | 743        | -         |          | 0.2 SFAM01.0 | PPB   | 10/28/2020 12:14 | Std  |
|           | 2.5 S2.5     | 8522       | -         |          | 2.5 SFAM01.0 | PPB   | 10/28/2020 12:16 | Std  |
|           | 5 S5.0       | 17684      | -         |          | 5 SFAM01.0   | PPB   | 10/28/2020 12:19 | Std  |
|           | 7.5 S7.5     | 26362      | -         |          | 7.5 SFAM01.0 | PPB   | 10/28/2020 12:21 | Std  |
|           | 10 S10.0     | 34745      | -         |          | 10 SFAM01.0  | PPB   | 10/28/2020 12:23 | Std  |
| ICV       | ICV          | 14205      | 4.0646 -  |          | SFAM01.0     | PPB   | 10/28/2020 12:26 | SMPL |
| ICB       | ICB          | -131       | -0.0421 - |          | SFAM01.0     | PPB   | 10/28/2020 12:28 | SMPL |
| CCV016    | CCV016       | 17540      | 5.0199 -  |          | SFAM01.0     | PPB   | 10/28/2020 12:30 | SMPL |
| CCB016    | CCB016       | -141       | -0.045 -  |          | SFAM01.0     | PPB   | 10/28/2020 12:33 | SMPL |
| PBW008    | PBW008       | 0          | -0.0046 - |          | SFAM01.0     | PPB   | 10/28/2020 12:35 | SMPL |
| PBW009    | PBW009       | 7          | -0.0026 - |          | SFAM01.0     | PPB   | 10/28/2020 12:37 | SMPL |
| L4485-08  | MDL-WATER-01 | 225        | 0.0599 -  |          | SFAM01.0     | PPB   | 10/28/2020 12:39 | SMPL |
| L4485-09  | MDL-WATER-02 | 201        | 0.053 -   |          | SFAM01.0     | PPB   | 10/28/2020 12:42 | SMPL |
| CCV017    | CCV017       | 18051      | 5.1663 -  |          | SFAM01.0     | PPB   | 10/28/2020 12:44 | SMPL |
| CCB017    | CCB017       | -108       | -0.0355 - |          | SFAM01.0     | PPB   | 10/28/2020 12:46 | SMPL |