



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB111723

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP87871
Turbidity Calibration std, 20NTU	WP87867
Turbidity Calibration std, 40NTU	WP87866
Turbidity Calibration std, 1NTU	WP87870
Turbidity Calibration std, 5NTU	WP87869
Turbidity Calibration - CCV std, 10 NTU	WP87868
10 NTU Standard 500 ml	W2752

Intercept: 0.1656

Slope: 0.9733

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.265	10/28/2020	12:05
2	CAL2	1	1	1.190	10/28/2020	12:07
3	CAL3	5	1	5.090	10/28/2020	12:09
4	CAL4	10	1	9.720	10/28/2020	12:11
5	CAL5	20	1	19.500	10/28/2020	12:13
6	CAL6	40	1	39.200	10/28/2020	12:15



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: Alexander

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB111723

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	10.100	10.207	10/28/2020	12:18
2	ICB		1	0.282	0.120	10/28/2020	12:20
3	CCV1	10	1	9.890	9.991	10/28/2020	12:22
4	CCB1		1	0.284	0.122	10/28/2020	12:24
5	LB111723BL		1	0.288	0.126	10/28/2020	12:27
6	L4542-01		1	9.890	9.991	10/28/2020	12:29
7	L4542-02		1	15.000	15.241	10/28/2020	12:32
8	L4542-02DUP		1	15.300	15.550	10/28/2020	12:35
9	CCV2	10	1	9.920	10.022	10/28/2020	12:38
10	CCB2		1	0.299	0.137	10/28/2020	12:40

LBW723

WORKLIST(Hardcopy Internal Chain)

WorkList Name : Turbidity-102820 WorkList ID : 143697 Department : Wet-Chemistry Date : 10-28-2020 10:44:04

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/29/2020	Water	L4542-01	Turbidity	Cool 4 deg C	LIRO01	D61	UDS-4-40	10/27/2020	SM2130 B
10/29/2020	Water	L4542-02	Turbidity	Cool 4 deg C	LIRO01	D61	UDS-6-16A	10/27/2020	SM2130 B

Date/Time 10/28/2020 11:40
Received by: AM (wet-chem)
Relinquished by: JP (wet-chem)

Date/Time 10/28/2020 12:50
Received by: JP (wet-chem)
Relinquished by: AM (wet-chem)