



PERCENT SOLID

Supervisor: apatel
Analyst: jignesh
Date: 10/29/2020

OVENTEMP IN Celsius(°C): 107
Time IN: 14:40
In Date: 10/28/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:02
Out Date: 10/29/2020
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB111725

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
L4545-01	MGB0B2	1	1.13	9.62	9.19	94.9	
L4545-02	MGB0B3	2	1.14	9.78	9.47	96.4	
L4545-03	MGB0B4	3	1.17	9.79	9.51	96.8	
L4545-04	MGB0B5	4	1.13	9.89	9.55	96.1	
L4545-05	MGB0B6	5	1.14	9.61	9.33	96.7	
L4545-06	MGB0B7	6	1.13	9.65	9.41	97.2	
L4545-07	MGB0B8	7	1.12	9.58	9.35	97.3	
L4545-08	MGB0B9	8	1.14	9.68	9.48	97.7	
L4545-09	MGB0C0	9	1.13	9.51	9.34	98.0	
L4545-10	MGB0C1	10	1.12	9.7	9.47	97.3	
L4545-11	MGB0C2	11	1.11	9.56	9.33	97.3	
L4545-12	MGB0C3	12	1.12	9.5	9.3	97.6	
L4545-13	MGB0C4	13	1.11	9.68	9.54	98.4	
L4545-14	MGB0C5	14	1.13	9.75	9.56	97.8	
L4545-15	MGB0C6	15	1.13	9.92	9.72	97.7	
L4545-16	MGB0C6D	16	1.13	9.92	9.72	97.7	
L4545-17	MGB0C6S	17	1.13	9.92	9.72	97.7	
L4545-18	MGB0C7	18	1.15	9.61	9.43	97.9	
L4545-19	MGB0C8	19	1.00	2.00	2.00	100.0	P.E. SAMPLE
L4545-20	MGB0C9	20	1.13	9.51	9.33	97.9	
L4545-21	MGB0D0	21	1.13	9.6	9.44	98.1	
L4545-22	MGB0F0	22	1.13	9.88	9.55	96.2	
L4545-23	MGB0F1	23	1.12	9.58	8.95	92.6	
L4545-24	MGB0F2	24	1.13	9.62	9.25	95.6	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$