


LB111727

7471B

INSTRUMENT ID: CV2

Linear



A= 0.0000e+000

B= 1.9367e-004

C= -2.3118e-002

Rho= 0.9998467

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.018	-0.018	28	0.000	28				
0.2	0.200	0.189	-0.011	1093	0.0 %	1093				
2.5	2.500	2.512	0.012	13090	0.0 %	13090				
5.0	5.000	4.982	-0.018	25842	0.0 %	25842				
7.5	7.500	7.626	0.126	39496	0.0 %	39496				
10.0	10.000	9.909	-0.091	51285	0.0 %	51285				

LB111727 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 28 Oct 2020 13:47:29

Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	Sc	28	-	0.00007471B	PPB	28 Oct 2020 13:48:40	S	Std
0.2 - 1	50.2	1093	-	0.20007471B	PPB	28 Oct 2020 13:50:57	S	Std
2.5 - 1	52.5	13090	-	2.50007471B	PPB	28 Oct 2020 13:53:14	S	Std
5.0 - 1	55	25842	-	5.00007471B	PPB	28 Oct 2020 13:55:31	S	Std
7.5 - 1	57.5	39496	-	7.50007471B	PPB	28 Oct 2020 13:57:47	S	Std
10.0 - 1	510	51285	-	10.00007471B	PPB	28 Oct 2020 14:00:03	S	Std
ICV77 - 1	ICV77	20118	3.8731	-7471B	PPB	28 Oct 2020 14:02:50	U	SMPL
ICB77 - 1	ICB77	-111	-0.0446	-7471B	PPB	28 Oct 2020 14:05:05	U	SMPL
CCV18 - 1	CCV18	25720	4.9581	-7471B	PPB	28 Oct 2020 14:07:20	U	SMPL
CCB18 - 1	CCB18	-104	-0.0433	-7471B	PPB	28 Oct 2020 14:09:35	U	SMPL
CRA - 1	CRA	1030	0.1764	-7471B	PPB	28 Oct 2020 14:11:50	U	SMPL
HighStd - 1	HighStd	50284	9.7154	-7471B	PPB	28 Oct 2020 14:14:05	U	SMPL
ChkStd - 1	ChkStd	37132	7.1683	-7471B	PPB	28 Oct 2020 14:21:37	U	SMPL
PB132666BL - 1	PBS	-136	-0.0495	-7471B	PPB	28 Oct 2020 14:23:52	U	SMPL
PB132666BS - 1	LCSS	18965	3.6498	-7471B	PPB	28 Oct 2020 14:26:09	U	SMPL
L4224-11 - 1	OR-02-102720	2013	0.3667	-7471B	PPB	28 Oct 2020 14:28:26	U	SMPL
L4224-11DUP - 1	OR-02-102720DUP	1940	0.3526	-7471B	PPB	28 Oct 2020 14:30:42	U	SMPL
L4224-11MS - 1	OR-02-102720MS	21613	4.1627	-7471B	PPB	28 Oct 2020 14:33:00	U	SMPL
L4224-11MSD - 1	OR-02-102720MSD	21571	4.1546	-7471B	PPB	28 Oct 2020 14:35:16	U	SMPL
L4523-05 - 1	CS-41	488	0.0714	-7471B	PPB	28 Oct 2020 14:37:32	U	SMPL
CCV19 - 1	CCV19	26115	5.0346	-7471B	PPB	28 Oct 2020 14:39:47	U	SMPL
CCB19 - 1	CCB19	-60	-0.0347	-7471B	PPB	28 Oct 2020 14:42:03	U	SMPL
L4523-06 - 1	CS-42	234	0.0222	-7471B	PPB	28 Oct 2020 14:44:18	U	SMPL
L4526-01 - 1	VNJ-225	3546	0.6636	-7471B	PPB	28 Oct 2020 14:46:34	U	SMPL
L4527-01 - 1	26414-26215	8737	1.6690	-7471B	PPB	28 Oct 2020 14:48:50	U	SMPL
L4527-03 - 1	VNJ-206	420	0.0582	-7471B	PPB	28 Oct 2020 14:51:05	U	SMPL
L4527-05 - 1	VNJ-232	1584	0.2837	-7471B	PPB	28 Oct 2020 14:53:21	U	SMPL
L4528-01 - 1	CHRT-27339	12785	2.4530	-7471B	PPB	28 Oct 2020 14:55:37	U	SMPL
L4528-03 - 1	CHRT-24245	8902	1.7009	-7471B	PPB	28 Oct 2020 14:57:53	U	SMPL
L4529-02 - 1	TP-T3	3008	0.5594	-7471B	PPB	28 Oct 2020 15:00:10	U	SMPL
L4530-02 - 1	RBR-54	1282	0.2252	-7471B	PPB	28 Oct 2020 15:02:27	U	SMPL
L4530-04 - 1	124021	1255	0.2199	-7471B	PPB	28 Oct 2020 15:04:43	U	SMPL
CCV20 - 1	CCV20	26011	5.0144	-7471B	PPB	28 Oct 2020 15:06:58	U	SMPL
CCB20 - 1	CCB20	-52	-0.0332	-7471B	PPB	28 Oct 2020 15:09:14	U	SMPL
L4537-01 - 1	S-1	1562	0.2794	-7471B	PPB	28 Oct 2020 15:11:29	U	SMPL
L4224-11LX5 - 1		367	0.0480	-7471B	PPB	28 Oct 2020 15:13:45	U	SMPL
L4224-11A - 1		21907	4.2196	-7471B	PPB	28 Oct 2020 15:16:00	U	SMPL
CCV21 - 1	CCV21	26254	5.0615	-7471B	PPB	28 Oct 2020 15:18:16	U	SMPL
CCB21 - 1	CCB21	-65	-0.0357	-7471B	PPB	28 Oct 2020 15:20:31	U	SMPL