

LB 111731


INSTRUMENT ID: CV1

ISM02.4

SFAM01.1

MB
10/28/20

Linear



A= 0.0000e+000

B= 2.6652e-004

C= -8.7160e-003

Rho= 0.9999537

Accept=Accepted

slope
+ intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.005	0.005	53	0.000	53					=
0.05	0.050					400					2
0.20	0.200	0.203	0.003	793	0.0 %	793					-1
2.50	2.500	2.471	-0.029	9304	0.0 %	9304					-1
5.00	5.000	5.057	0.057	19006	0.0 %	19006					-1
7.50	7.500	7.445	-0.055	27965	0.0 %	27965					-1
10.0	10.000	10.020	0.020	37627	0.0 %	37627					0

LB111731

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	53	-		0 SFAM01.0	PPB	10/28/2020 15:30	Std
	0.2 S0.2	793	-		0.2 SFAM01.0	PPB	10/28/2020 15:33	Std
	2.5 S2.5	9304	-		2.5 SFAM01.0	PPB	10/28/2020 15:35	Std
	5 S5.0	19006	-		5 SFAM01.0	PPB	10/28/2020 15:37	Std
	7.5 S7.5	27965	-		7.5 SFAM01.0	PPB	10/28/2020 15:39	Std
	10 S10.0	37627	-		10 SFAM01.0	PPB	10/28/2020 15:42	Std
ICV78	ICV78	14972	3.9816 -		SFAM01.0	PPB	10/28/2020 15:44	SMPL
ICB78	ICB78	0	-0.0087 -		SFAM01.0	PPB	10/28/2020 15:47	SMPL
ICB78	ICB78	-49	-0.0218 -		SFAM01.0	PPB	10/28/2020 15:48	SMPL
CCV22	CCV22	19077	5.0757 -		SFAM01.0	PPB	10/28/2020 15:50	SMPL
CCB22	CCB22	-113	-0.0388 -		SFAM01.0	PPB	10/28/2020 15:53	SMPL
PBS003	PBS003	-22	-0.0146 -		SFAM01.0	PPB	10/28/2020 15:55	SMPL
PBS004	PBS004	-32	-0.0172 -		SFAM01.0	PPB	10/28/2020 15:57	SMPL
L4485-03	MDL-SOIL-03	215	0.0486 -		SFAM01.0	PPB	10/28/2020 15:59	SMPL
L4485-04	MDL-SOIL-04	223	0.0507 -		SFAM01.0	PPB	10/28/2020 16:04	SMPL
CCV23	CCV23	18809	5.0043 -		SFAM01.0	PPB	10/28/2020 16:06	SMPL
CCB23	CCB23	-99	-0.0351 -		SFAM01.0	PPB	10/28/2020 16:09	SMPL