



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111739

Slope : 98.4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/28/2020	16:25
2	CAL2	1	Water	NA	NA	20.1	7.00	10/28/2020	16:26
3	CAL3	1	Water	NA	NA	20.1	10.03	10/28/2020	16:27
4	ICV	1	Water	NA	NA	20.1	7.00	10/28/2020	16:29
5	CCV1	1	Water	NA	NA	20.2	2.01	10/28/2020	16:30
6	L4539-04	1	Water	NA	NA	13.0	7.80	10/28/2020	16:32
7	L4539-04DUP	1	Water	NA	NA	13.1	7.81	10/28/2020	16:33
8	CCV2	1	Water	NA	NA	20.1	12.02	10/28/2020	16:35

VB 111739

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4539 WorkList ID : 143718 Department : Wet-Chemistry Date : 10-28-2020 16:18:22

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/28/2020	Water	L4539-04	pH	Cool 4 deg C	PSEG01	D61	FM655	10/28/2020	9040C

Date/Time 10-28-2020 16:20
Received by: JN (wet Chem)
Relinquished by: CP (SM)

Date/Time 10-28-2020 17:20
Received by: CP (SM)
Relinquished by: JN (wet Chem)