


LB111742

SFAM01.0

INSTRUMENT ID: CV1

Linear



A= 0.0000e+000

B= 2.6418e-004

C= 2.4135e-002

Rho= 0.9998337

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%AD
0.00	0.000	0.027	0.027	12	0.000	12					—
0.05	0.050					400					16
0.20	0.200	0.232	0.032	785	0.0 %	785					-3
2.50	2.500	2.434	-0.066	9122	0.0 %	9122					1
5.00	5.000	5.045	0.045	19007	0.0 %	19007					-2
7.50	7.500	7.385	-0.115	27864	0.0 %	27864					1
10.0	10.000	10.076	0.076	38051	0.0 %	38051					

LB111742

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	12	-		0 SFAM01.0	PPB	10/29/2020 12:40	Std
	0.2 S0.2	785	-		0.2 SFAM01.0	PPB	10/29/2020 12:42	Std
	2.5 S2.5	9122	-		2.5 SFAM01.0	PPB	10/29/2020 12:44	Std
	5 S5.0	19007	-		5 SFAM01.0	PPB	10/29/2020 12:46	Std
	7.5 S7.5	27864	-		7.5 SFAM01.0	PPB	10/29/2020 12:49	Std
	10 S10.0	38051	-		10 SFAM01.0	PPB	10/29/2020 12:51	Std
ICV	ICV	15067	4.0046 -		SFAM01.0	PPB	10/29/2020 12:54	SMPL
ICB	ICB	-116	-0.0065 -		SFAM01.0	PPB	10/29/2020 12:56	SMPL
CCV028	CCV028	18567	4.9292 -		SFAM01.0	PPB	10/29/2020 12:58	SMPL
CCB028	CCB028	-139	-0.0126 -		SFAM01.0	PPB	10/29/2020 13:00	SMPL
PBS005	PBS005	-40	0.0136 -		SFAM01.0	PPB	10/29/2020 13:03	SMPL
PBS006	PBS006	-34	0.0152 -		SFAM01.0	PPB	10/29/2020 13:05	SMPL
PBS007	PBS007	10	0.0268 -		SFAM01.0	PPB	10/29/2020 13:11	SMPL
L4485-05	MDL-SOIL-05	162	0.0669 -		SFAM01.0	PPB	10/29/2020 13:13	SMPL
L4485-06	MDL-SOIL-06	171	0.0693 -		SFAM01.0	PPB	10/29/2020 13:16	SMPL
L4485-07	MDL-SOIL-07	105	0.0519 -		SFAM01.0	PPB	10/29/2020 13:18	SMPL
CCV029	CCV029	18794	4.9892 -		SFAM01.0	PPB	10/29/2020 13:20	SMPL
CCB029	CCB029	-90	0.0004 -		SFAM01.0	PPB	10/29/2020 13:23	SMPL