


LB111745

SFAM01.0

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.9076e-004

C= -2.0933e-003

Rho= 0.9998983

Accept=Accepted

slope
y=1.51e-004x

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	% D
0.00	0.000	0.011	0.011	45	0.000	45					1
0.05	0.050					400					1
0.20	0.200	0.222	0.022	770	0.0 %	770					1
2.50	2.500	2.418	-0.082	8324	0.0 %	8324					-3
5.00	5.000	5.022	0.022	17279	0.0 %	17279					0
7.50	7.500	7.579	0.079	26072	0.0 %	26072					1
10.0	10.000	9.948	-0.052	34222	0.0 %	34222					-1

LB111745

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	45	-		0 SFAM01.0	PPB	10/29/2020 14:06	Std
	0.2 S0.2	770	-		0.2 SFAM01.0	PPB	10/29/2020 14:09	Std
	2.5 S2.5	8324	-		2.5 SFAM01.0	PPB	10/29/2020 14:11	Std
	5 S5.0	17279	-		5 SFAM01.0	PPB	10/29/2020 14:13	Std
	7.5 S7.5	26072	-		7.5 SFAM01.0	PPB	10/29/2020 14:17	Std
	10 S10.0	34222	-		10 SFAM01.0	PPB	10/29/2020 14:19	Std
ICV	ICV	14146	4.111 -		SFAM01.0	PPB	10/29/2020 14:22	SMPL
ICB	ICB	-133	-0.0408 -		SFAM01.0	PPB	10/29/2020 14:24	SMPL
CCV030	CCV030	17461	5.0749 -		SFAM01.0	PPB	10/29/2020 14:26	SMPL
CCB030	CCB030	-138	-0.0422 -		SFAM01.0	PPB	10/29/2020 14:29	SMPL
PBW010	PBW010	133	0.0366 -		SFAM01.0	PPB	10/29/2020 14:31	SMPL
PBW011	PBW011	61	0.0156 -		SFAM01.0	PPB	10/29/2020 14:33	SMPL
L4485-10	MDL-WATER-03	242	0.0683 -		SFAM01.0	PPB	10/29/2020 14:35	SMPL
L4485-11	MDL-WATER-04	219	0.0616 -		SFAM01.0	PPB	10/29/2020 14:40	SMPL
CCV031	CCV031	17328	5.0362 -		SFAM01.0	PPB	10/29/2020 14:42	SMPL
CCB031	CCB031	-88	-0.0277 -		SFAM01.0	PPB	10/29/2020 14:45	SMPL