



Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB111752

Slope : 98.1

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2629
buffer solution pH 7 yellow	W2627
PH 10.01 BUFFER,COLOR CD 475ML	W2632
buffer solution pH 7 yellow	W2569
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2616

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/29/2020	16:10
2	CAL2	1	Water	NA	NA	20.1	6.99	10/29/2020	16:11
3	CAL3	1	Water	NA	NA	20.1	10.03	10/29/2020	16:12
4	ICV	1	Water	NA	NA	20.2	7.00	10/29/2020	16:14
5	CCV1	1	Water	NA	NA	20.1	2.01	10/29/2020	16:15
6	L4553-01	1	Solid	20.03	20	20.3	8.95	10/29/2020	16:16
7	L4553-03	1	Solid	20.04	20	19.6	7.87	10/29/2020	16:19
8	L4553-05	1	Solid	20.02	20	19.3	8.50	10/29/2020	16:20
9	L4553-05DUP	1	Solid	20.03	20	19.4	8.51	10/29/2020	16:21
10	CCV2	1	Water	NA	NA	20.1	12.02	10/29/2020	16:22

11752

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH L4553 WorkList ID : 143758 Department : Wet-Chemistry Date : 10-29-2020 15:51:11

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/29/2020	Solid	L4553-01	pH	Cool 4 deg C	PSEG01	E11	COMP-1	10/29/2020	9045D
10/29/2020	Solid	L4553-03	pH	Cool 4 deg C	PSEG01	E11	COMP-2	10/29/2020	9045D
10/29/2020	Solid	L4553-05	pH	Cool 4 deg C	PSEG01	E11	CHRT-26601	10/29/2020	9045D

Date/Time 10-29-2020 16:00
 Received by: [Signature]
 Relinquished by: [Signature]

Date/Time 10-29-2020 17:00
 Received by: [Signature]
 Relinquished by: [Signature]